

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060723\
 Data File : VN078104.D
 Acq On : 07 Jun 2023 15:28
 Operator : JC\MD
 Sample : 03037-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Jun 08 05:28:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	383231	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	707685	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	609236	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	180138	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	260203	50.924	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.840%
35) Dibromofluoromethane	8.171	113	226906	50.710	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.420%
50) Toluene-d8	10.571	98	853695	49.489	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.980%
62) 4-Bromofluorobenzene	12.853	95	240531	44.853	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060723\
Data File : VN078104.D
Acq On : 07 Jun 2023 15:28
Operator : JC\MD
Sample : 03037-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Jun 08 05:28:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

