

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060723\
 Data File : VN078111.D
 Acq On : 07 Jun 2023 18:15
 Operator : JC\MD
 Sample : 03046-01 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-1-1-5

Quant Time: Jun 08 05:29:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	374576	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	695502	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	618239	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	188317	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	257399	51.539	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.080%
35) Dibromofluoromethane	8.177	113	224488	51.048	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.100%
50) Toluene-d8	10.571	98	848873	50.072	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.140%
62) 4-Bromofluorobenzene	12.853	95	255969	48.568	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.140%
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	58749	40.868	ug/l	89
18) Methyl Acetate	5.041	43	11397	1.722	ug/l #	64
20) Methylene Chloride	5.289	84	39238	8.260	ug/l	91
25) 2-Butanone	7.494	43	47887	21.662	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060723\
Data File : VN078111.D
Acq On : 07 Jun 2023 18:15
Operator : JC\MD
Sample : 03046-01 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-1-1-5

Quant Time: Jun 08 05:29:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

