

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060723\
 Data File : VN078112.D
 Acq On : 07 Jun 2023 18:39
 Operator : JC\MD
 Sample : 03046-07 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-2-6-10

Quant Time: Jun 08 05:29:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/08/2023
 Supervised By : Mahesh Dadoda 06/08/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	373511	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	689108	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	612184	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	184935	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	256227	51.451	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 102.900%			
35) Dibromofluoromethane	8.177	113	222836	51.143	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 102.280%			
50) Toluene-d8	10.571	98	846507	50.396	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.800%			
62) 4-Bromofluorobenzene	12.853	95	249210	47.725	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 95.440%			
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	16066	11.208	ug/l	91
18) Methyl Acetate	5.053	43	7904m	1.198	ug/l	
20) Methylene Chloride	5.283	84	21801	4.602	ug/l #	92
25) 2-Butanone	7.489	43	46139	20.931	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060723\
Data File : VN078112.D
Acq On : 07 Jun 2023 18:39
Operator : JC\MD
Sample : 03046-07 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-2-6-10

Quant Time: Jun 08 05:29:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 06/08/2023
Supervised By : Mahesh Dadoda 06/08/2023

