



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060723\  
Data File : VN078088.D  
Acq On : 07 Jun 2023 08:37  
Operator : JC\MD  
Sample : BFB  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060623W.M  
Title : SW846 8260  
Last Update : Wed Jun 07 05:05:00 2023



AutoFind: Scans 1852, 1853, 1854; Background Corrected with Scan 1844

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 50             | 95              | 15              | 40              | 20.0         | 11295      | PASS                |
| 75             | 95              | 30              | 60              | 56.8         | 32144      | PASS                |
| 95             | 95              | 100             | 100             | 100.0        | 56603      | PASS                |
| 96             | 95              | 5               | 9               | 7.5          | 4267       | PASS                |
| 173            | 174             | 0.00            | 2               | 1.1          | 469        | PASS                |
| 174            | 95              | 50              | 100             | 74.4         | 42133      | PASS                |
| 175            | 174             | 5               | 9               | 7.2          | 3051       | PASS                |
| 176            | 174             | 95              | 101             | 100.4        | 42285      | PASS                |
| 177            | 176             | 5               | 9               | 7.2          | 3033       | PASS                |

Average of 12.847 to 12.859 min.: VN078088.D\data.ms

BFB

Modified:subtracted

| m/z   | abund. | m/z   | abund. | m/z   | abund. | m/z   | abund. |
|-------|--------|-------|--------|-------|--------|-------|--------|
| 35.85 | 408    | 48.10 | 156    | 61.10 | 888    | 73.05 | 296    |
| 36.20 | 350    | 49.00 | 2427   | 61.95 | 3206   | 74.00 | 1168   |
| 37.00 | 2961   | 50.00 | 11295  | 63.05 | 2495   | 75.00 | 3214   |
| 38.05 | 2404   | 51.05 | 3488   | 64.00 | 389    | 76.00 | 283    |
| 39.10 | 1033   | 51.90 | 170    | 65.00 | 55     | 76.90 | 466    |
| 41.00 | 99     | 55.10 | 270    | 66.90 | 101    | 77.80 | 266    |
| 44.00 | 449    | 55.80 | 325    | 68.00 | 6463   | 78.85 | 1997   |
| 44.90 | 615    | 56.00 | 757    | 69.05 | 6906   | 79.75 | 356    |
| 46.80 | 116    | 57.00 | 1835   | 69.80 | 82     | 80.10 | 324    |
| 47.05 | 468    | 60.00 | 500    | 70.10 | 559    | 80.95 | 2527   |
| 47.85 | 360    | 60.90 | 1912   | 71.95 | 504    | 81.85 | 489    |

Average of 12.847 to 12.859 min.: VN078088.D\data.ms

BFB

Modified:subtracted

| m/z   | abund. | m/z    | abund. | m/z    | abund. | m/z    | abund. |
|-------|--------|--------|--------|--------|--------|--------|--------|
| 86.10 | 50     | 102.90 | 61     | 118.85 | 431    | 150.00 | 52     |
| 86.95 | 2615   | 103.60 | 113    | 128.00 | 197    | 153.10 | 53     |
| 88.00 | 3065   | 103.90 | 202    | 129.80 | 87     | 155.00 | 85     |
| 89.10 | 50     | 105.10 | 53     | 130.95 | 144    | 156.90 | 106    |
| 90.85 | 343    | 105.85 | 334    | 135.00 | 61     | 171.30 | 50     |
| 92.05 | 1903   | 113.00 | 64     | 137.10 | 57     | 171.80 | 262    |
| 92.95 | 2793   | 115.10 | 121    | 140.90 | 687    | 172.60 | 113    |
| 94.00 | 7507   | 115.65 | 119    | 142.85 | 1067   | 172.90 | 469    |
| 95.00 | 56603  | 115.95 | 209    | 145.20 | 120    | 173.10 | 166    |
| 96.00 | 4267   | 116.95 | 619    | 147.80 | 52     | 174.00 | 42133  |
| 97.00 | 55     | 117.85 | 267    | 148.90 | 74     | 174.95 | 3051   |

Average of 12.847 to 12.859 min.: VN078088.D\data.ms

BFB

Modified:subtracted

| m/z    | abund. | m/z | abund. | m/z | abund. | m/z | abund. |
|--------|--------|-----|--------|-----|--------|-----|--------|
| 176.00 | 42285  |     |        |     |        |     |        |
| 176.95 | 3033   |     |        |     |        |     |        |
| 281.20 | 159    |     |        |     |        |     |        |

Instrument :  
MSVOA\_N  
ClientSampleId :  
BFB