

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060724\
 Data File : VN082486.D
 Acq On : 07 Jun 2024 12:35
 Operator : JC\MD
 Sample : P2771-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
ClientSampleId :
 BP-VPB-196-EB-20240604

Quant Time: Jun 08 00:56:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	166216	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	299981	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	287708	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	122178	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	136021	51.826	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.660%
35) Dibromofluoromethane	8.165	113	97647	52.915	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.840%
50) Toluene-d8	10.571	98	363079	51.319	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.640%
62) 4-Bromofluorobenzene	12.853	95	119420	47.291	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	94.580%
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	7125	6.618	ug/l	95
25) 2-Butanone	7.494	43	7438	4.557	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060724\
Data File : VN082486.D
Acq On : 07 Jun 2024 12:35
Operator : JC\MD
Sample : P2771-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-EB-20240604

Quant Time: Jun 08 00:56:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

