

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060724\
 Data File : VN082492.D
 Acq On : 07 Jun 2024 15:00
 Operator : JC\MD
 Sample : P2771-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-196-GW-100-102

Quant Time: Jun 08 00:58:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	161172	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	290120	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	284649	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	118031	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	133586	52.491	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 104.980%			
35) Dibromofluoromethane	8.171	113	99256	55.615	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 111.240%			
50) Toluene-d8	10.571	98	361444	52.825	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 105.640%			
62) 4-Bromofluorobenzene	12.853	95	118801	48.645	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery = 97.300%			
Target Compounds						
16) Acetone	4.436	43	5940	5.690	ug/l #	84
19) Methyl tert-butyl Ether	5.806	73	6523	1.014	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060724\
Data File : VN082492.D
Acq On : 07 Jun 2024 15:00
Operator : JC\MD
Sample : P2771-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-100-102

Quant Time: Jun 08 00:58:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

