

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060724\
 Data File : VN082493.D
 Acq On : 07 Jun 2024 15:24
 Operator : JC\MD
 Sample : P2771-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-196-GW-160-162

Quant Time: Jun 08 00:59:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	155440	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	290634	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	278262	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	116496	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	131519	53.584	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.160%	
35) Dibromofluoromethane	8.171	113	97658	54.623	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 109.240%	
50) Toluene-d8	10.571	98	351544	51.287	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.580%	
62) 4-Bromofluorobenzene	12.853	95	119486	48.839	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 97.680%	
Target Compounds						
16) Acetone	4.430	43	10058	9.990	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060724\
Data File : VN082493.D
Acq On : 07 Jun 2024 15:24
Operator : JC\MD
Sample : P2771-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-160-162

Quant Time: Jun 08 00:59:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

