

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060724\
 Data File : VN082496.D
 Acq On : 07 Jun 2024 16:38
 Operator : JC\MD
 Sample : P2771-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-196-DUP-20240604

Quant Time: Jun 08 01:00:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	154410	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	290629	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	279727	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	117974	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	132197	54.220	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.440%	
35) Dibromofluoromethane	8.171	113	96117	53.762	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.520%	
50) Toluene-d8	10.571	98	348558	50.852	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.700%	
62) 4-Bromofluorobenzene	12.853	95	116283	47.531	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 95.060%	
Target Compounds						
16) Acetone	4.447	43	5701	5.700	ug/l	Qvalue # 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060724\
Data File : VN082496.D
Acq On : 07 Jun 2024 16:38
Operator : JC\MD
Sample : P2771-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-DUP-20240604

Quant Time: Jun 08 01:00:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

