

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060724\
 Data File : VN082499.D
 Acq On : 07 Jun 2024 17:51
 Operator : JC\MD
 Sample : P2772-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-189-GW-58-60

Quant Time: Jun 08 01:01:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	156168	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	288454	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	281476	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	116616	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	130641	52.978	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 105.960%	
35) Dibromofluoromethane	8.165	113	93842	52.886	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.780%	
50) Toluene-d8	10.571	98	351972	51.737	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.480%	
62) 4-Bromofluorobenzene	12.853	95	116719	48.069	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 96.140%	
Target Compounds						
						Qvalue
16) Acetone	4.442	43	9106	9.002	ug/l #	86
30) Chloroform	7.965	83	4837	1.289	ug/l	89
64) Tetrachloroethene	11.112	164	2599	1.246	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060724\
Data File : VN082499.D
Acq On : 07 Jun 2024 17:51
Operator : JC\MD
Sample : P2772-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-189-GW-58-60

Quant Time: Jun 08 01:01:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

