

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060724\
 Data File : VN082479.D
 Acq On : 07 Jun 2024 09:25
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 08 00:52:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	92	0.00
2 T	Dichlorodifluoromethane	0.721	0.755	-4.7	93	0.00
3 P	Chloromethane	0.817	0.813	0.5	93	0.00
4 C	Vinyl Chloride	0.768	0.785	-2.2#	93	0.00
5 T	Bromomethane	0.441	0.420	4.8	93	0.00
6 T	Chloroethane	0.501	0.515	-2.8	99	0.00
7 T	Trichlorofluoromethane	1.006	1.062	-5.6	96	0.00
8 T	Diethyl Ether	0.358	0.355	0.8	87	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.588	0.608	-3.4	97	0.00
10 T	Methyl Iodide	0.591	0.648	-9.6	96	0.00
11 T	Tert butyl alcohol	0.127	0.111	12.6	79	0.00
12 CM	1,1-Dichloroethene	0.546	0.564	-3.3#	93	0.00
13 T	Acrolein	0.094	0.106	-12.8	98	0.00
14 T	Allyl chloride	1.169	1.110	5.0	90	0.00
15 T	Acrylonitrile	0.328	0.342	-4.3	93	0.00
16 T	Acetone	0.324	0.312	3.7	83	0.00
17 T	Carbon Disulfide	1.764	1.760	0.2	96	0.00
18 T	Methyl Acetate	0.807	0.853	-5.7	100	0.00
19 T	Methyl tert-butyl Ether	1.996	2.076	-4.0	92	0.00
20 T	Methylene Chloride	0.653	0.663	-1.5	97	0.00
21 T	trans-1,2-Dichloroethene	0.595	0.621	-4.4	97	0.00
22 T	Diisopropyl ether	2.441	2.506	-2.7	92	0.00
23 T	Vinyl Acetate	1.925	2.033	-5.6	91	0.00
24 P	1,1-Dichloroethane	1.258	1.292	-2.7	95	0.00
25 T	2-Butanone	0.491	0.498	-1.4	90	0.00
26 T	2,2-Dichloropropane	1.054	1.117	-6.0	95	0.00
27 T	cis-1,2-Dichloroethene	0.702	0.741	-5.6	96	0.00
28 T	Bromochloromethane	0.586	0.532	9.2	89	0.00
29 T	Tetrahydrofuran	0.315	0.338	-7.3	94	0.00
30 C	Chloroform	1.202	1.274	-6.0#	98	0.00
31 T	Cyclohexane	1.261	1.198	5.0	94	0.00
32 T	1,1,1-Trichloroethane	1.066	1.134	-6.4	96	0.00
33 S	1,2-Dichloroethane-d4	0.790	0.729	7.7	85	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.308	0.316	-2.6	90	0.00
36 T	1,1-Dichloropropene	0.516	0.569	-10.3	96	0.00
37 T	Ethyl Acetate	0.577	0.622	-7.8	90	0.00
38 T	Carbon Tetrachloride	0.523	0.601	-14.9	99	0.00
39 T	Methylcyclohexane	0.565	0.623	-10.3	92	0.00
40 TM	Benzene	1.538	1.718	-11.7	97	0.00
41 T	Methacrylonitrile	0.340	0.355	-4.4	91	0.00
42 TM	1,2-Dichloroethane	0.592	0.649	-9.6	97	0.00
43 T	Isopropyl Acetate	1.023	1.112	-8.7	93	0.00
44 TM	Trichloroethene	0.349	0.384	-10.0	97	0.00
45 C	1,2-Dichloropropane	0.397	0.441	-11.1#	97	0.00
46 T	Dibromomethane	0.261	0.290	-11.1	97	0.00
47 T	Bromodichloromethane	0.550	0.632	-14.9	98	0.00
48 T	Methyl methacrylate	0.487	0.535	-9.9	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060724\
 Data File : VN082479.D
 Acq On : 07 Jun 2024 09:25
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 08 00:52:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.008	-14.3	93	0.00
50 S	Toluene-d8	1.179	1.139	3.4	84	0.00
51 T	4-Methyl-2-Pentanone	0.581	0.652	-12.2	93	0.00
52 CM	Toluene	0.923	1.057	-14.5#	97	0.00
53 T	t-1,3-Dichloropropene	0.569	0.659	-15.8	95	0.00
54 T	cis-1,3-Dichloropropene	0.602	0.694	-15.3	96	0.00
55 T	1,1,2-Trichloroethane	0.335	0.382	-14.0	100	0.00
56 T	Ethyl methacrylate	0.570	0.657	-15.3	92	0.00
57 T	1,3-Dichloropropane	0.627	0.690	-10.0	96	0.00
58 T	2-Chloroethyl Vinyl ether	0.294	0.316	-7.5	88	0.00
59 T	2-Hexanone	0.426	0.488	-14.6	92	0.00
60 T	Dibromochloromethane	0.370	0.464	-25.4#	101	0.00
61 T	1,2-Dibromoethane	0.345	0.395	-14.5	100	0.00
62 S	4-Bromofluorobenzene	0.421	0.435	-3.3	88	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	91	0.00
64 T	Tetrachloroethene	0.371	0.419	-12.9	104	0.00
65 PM	Chlorobenzene	1.105	1.215	-10.0	99	0.00
66 T	1,1,1,2-Tetrachloroethane	0.364	0.425	-16.8	101	0.00
67 C	Ethyl Benzene	1.976	2.174	-10.0#	95	0.00
68 T	m/p-Xylenes	0.724	0.825	-14.0	97	0.00
69 T	o-Xylene	0.701	0.780	-11.3	95	0.00
70 T	Styrene	1.165	1.354	-16.2	97	0.00
71 P	Bromoform	0.261	0.334	-28.0#	104	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
73 T	Isopropylbenzene	3.972	3.999	-0.7	94	0.00
74 T	N-amyl acetate	2.185	2.078	4.9	90	0.00
75 P	1,1,2,2-Tetrachloroethane	1.145	1.135	0.9	97	0.00
76 T	1,2,3-Trichloropropane	1.052	0.933	11.3	84	0.00
77 T	Bromobenzene	0.947	0.973	-2.7	100	0.00
78 T	n-propylbenzene	4.704	4.930	-4.8	96	0.00
79 T	2-Chlorotoluene	2.954	2.961	-0.2	96	0.00
80 T	1,3,5-Trimethylbenzene	3.212	3.417	-6.4	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.450	0.467	-3.8	98	0.00
82 T	4-Chlorotoluene	2.888	2.987	-3.4	98	0.00
83 T	tert-Butylbenzene	2.810	2.808	0.1	93	0.00
84 T	1,2,4-Trimethylbenzene	3.286	3.438	-4.6	95	0.00
85 T	sec-Butylbenzene	3.895	4.025	-3.3	95	0.00
86 T	p-Isopropyltoluene	3.165	3.397	-7.3	96	0.00
87 T	1,3-Dichlorobenzene	1.715	1.795	-4.7	100	0.00
88 T	1,4-Dichlorobenzene	1.750	1.803	-3.0	101	0.00
89 T	n-Butylbenzene	2.845	3.016	-6.0	95	0.00
90 T	Hexachloroethane	0.586	0.633	-8.0	99	0.00
91 T	1,2-Dichlorobenzene	1.623	1.714	-5.6	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.235	0.248	-5.5	95	0.00
93 T	1,2,4-Trichlorobenzene	0.910	0.916	-0.7	93	0.00
94 T	Hexachlorobutadiene	0.399	0.395	1.0	97	0.00
95 T	Naphthalene	2.852	3.011	-5.6	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060724\
Data File : VN082479.D
Acq On : 07 Jun 2024 09:25
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jun 08 00:52:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.911	0.934	-2.5	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6