

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060819\
 Data File : VN056086.D
 Acq On : 7 Jun 2019 9:19
 Operator : JC/SP
 Sample : VN0608WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0608WBL01

Quant Time: Jun 07 10:03:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N060619W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jun 06 04:02:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.19	128	49566	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.58	114	257734	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	198976	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.02	65	95569	31.05	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	103.50%
60) 4-Bromofluorobenzene	12.40	95	63600	21.56	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	71.87%
63) Toluene-d8	10.09	98	297877	33.00	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	110.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN060819\
Data File : VN056086.D
Acq On : 7 Jun 2019 9:19
Operator : JC/SP
Sample : VN0608WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0608WBL01

Quant Time: Jun 07 10:03:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N060619W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Jun 06 04:02:53 2019
Response via : Initial Calibration

