

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060819\
 Data File : VN056087.D
 Acq On : 7 Jun 2019 9:45
 Operator : JC/SP
 Sample : K3232-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPDES-1

Quant Time: Jun 07 10:07:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N060619W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jun 06 04:02:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	50243	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.58	114	263945	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	209976	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.02	65	95843	30.72	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	102.40%
60) 4-Bromofluorobenzene	12.40	95	69108	22.20	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	74.00%
63) Toluene-d8	10.09	98	308649	32.40	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	108.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN060819\
Data File : VN056087.D
Acq On : 7 Jun 2019 9:45
Operator : JC/SP
Sample : K3232-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPDES-1

Quant Time: Jun 07 10:07:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N060619W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Jun 06 04:02:53 2019
Response via : Initial Calibration

