

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060819\
 Data File : VN056104.D
 Acq On : 7 Jun 2019 17:39
 Operator : JC/SP
 Sample : K3223-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Z1-TB11-190605

Quant Time: Jun 08 06:30:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	334145	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	501944	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	470003	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	148136	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	169364	51.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.10%	
35) Dibromofluoromethane	7.59	113	153189	49.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.48%	
50) Toluene-d8	10.09	98	629735	53.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.28%	
62) 4-Bromofluorobenzene	12.40	95	196286	48.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.42%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN060819\
Data File : VN056104.D
Acq On : 7 Jun 2019 17:39
Operator : JC/SP
Sample : K3223-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Z1-TB11-190605

Quant Time: Jun 08 06:30:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

