

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060819\
 Data File : VN056106.D
 Acq On : 7 Jun 2019 18:30
 Operator : JC/SP
 Sample : K3251-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Jun 08 06:35:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	325422	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	489237	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	458359	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	144725	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	159564	49.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.76%	
35) Dibromofluoromethane	7.59	113	139879	46.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.20%	
50) Toluene-d8	10.09	98	589377	51.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.06%	
62) 4-Bromofluorobenzene	12.40	95	182438	46.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.90%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN060819\
Data File : VN056106.D
Acq On : 7 Jun 2019 18:30
Operator : JC/SP
Sample : K3251-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Jun 08 06:35:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

