

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060819\
 Data File : VN056115.D
 Acq On : 7 Jun 2019 22:15
 Operator : JC/SP
 Sample : K3203-09DL 100X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 LF015MW045-190604DL

Quant Time: Jun 08 06:58:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	331549	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	498789	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	476058	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	155323	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	171608	52.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.26%	
35) Dibromofluoromethane	7.59	113	151505	49.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.02%	
50) Toluene-d8	10.09	98	629632	53.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.94%	
62) 4-Bromofluorobenzene	12.40	95	200162	49.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.96%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	2.18	62	143145	39.706	ug/l	100
27) cis-1,2-Dichloroethene	6.83	96	11561	3.298	ug/l	75
40) Benzene	8.04	78	11907	0.914	ug/l	93
65) Chlorobenzene	11.43	112	195780	20.664	ug/l	100
87) 1,3-Dichlorobenzene	13.28	146	2100	0.408	ug/l	97
88) 1,4-Dichlorobenzene	13.36	146	16523	3.337	ug/l	95
91) 1,2-Dichlorobenzene	13.65	146	48351	9.305	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN060819\
Data File : VN056115.D
Acq On : 7 Jun 2019 22:15
Operator : JC/SP
Sample : K3203-09DL 100X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
LF015MW045-190604DL

Quant Time: Jun 08 06:58:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

