

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060819\
 Data File : VN056135.D
 Acq On : 8 Jun 2019 7:27
 Operator : JC/SP
 Sample : K3214-03 50X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-4A

Quant Time: Jun 10 01:57:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	311148	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	474721	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	453207	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	167039	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	154983	50.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.34%	
35) Dibromofluoromethane	7.59	113	136435	46.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.68%	
50) Toluene-d8	10.09	98	572114	51.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.10%	
62) 4-Bromofluorobenzene	12.40	95	189057	49.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.20%	
Target Compounds						
40) Benzene	8.04	78	16927	1.365	ug/l	92
67) Ethyl Benzene	11.51	91	15100	0.963	ug/l	97
68) m/p-Xylenes	11.62	106	3932	0.656	ug/l	99
69) o-Xylene	11.95	106	3035	0.515	ug/l	100
80) 1,3,5-Trimethylbenzene	12.73	105	4936	0.439	ug/l	98
84) 1,2,4-Trimethylbenzene	13.04	105	14413	1.353	ug/l	94
86) p-Isopropyltoluene	13.29	119	3449	0.308	ug/l	95
95) Naphthalene	15.13	128	186216	25.363	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN060819\
Data File : VN056135.D
Acq On : 8 Jun 2019 7:27
Operator : JC/SP
Sample : K3214-03 50X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-4A

Quant Time: Jun 10 01:57:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

