

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060819\
 Data File : VN056139.D
 Acq On : 8 Jun 2019 9:07
 Operator : JC/SP
 Sample : K3160-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 58 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EO-02-060619-B

Quant Time: Jun 10 02:12:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	321714	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	489312	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	476943	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	156535	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	161597	50.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.18%	
35) Dibromofluoromethane	7.59	113	140984	46.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.92%	
50) Toluene-d8	10.09	98	598431	51.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.60%	
62) 4-Bromofluorobenzene	12.40	95	191771	48.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.62%	
Target Compounds						
16) Acetone	3.82	43	19446	12.728	ug/l	100
20) Methylene Chloride	4.55	84	98360	31.250	ug/l	97
43) Isopropyl Acetate	8.17	43	8663	1.428	ug/l #	82
95) Naphthalene	15.13	128	14710	5.665	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN060819\
Data File : VN056139.D
Acq On : 8 Jun 2019 9:07
Operator : JC/SP
Sample : K3160-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 58 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EO-02-060619-B

Quant Time: Jun 10 02:12:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

