

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060819\
 Data File : VN056141.D
 Acq On : 8 Jun 2019 9:57
 Operator : JC/SP
 Sample : K3156-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 60 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HD-02-060519

Quant Time: Jun 10 02:16:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	325710	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	502250	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	484455	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	157936	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	164488	50.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.72%	
35) Dibromofluoromethane	7.59	113	143287	46.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.00%	
50) Toluene-d8	10.09	98	605267	51.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.10%	
62) 4-Bromofluorobenzene	12.40	95	196260	48.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.34%	
Target Compounds						
16) Acetone	3.82	43	8574	5.543	ug/l	98
20) Methylene Chloride	4.54	84	279000	87.553	ug/l	98
43) Isopropyl Acetate	8.17	43	10171	1.633	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN060819\
Data File : VN056141.D
Acq On : 8 Jun 2019 9:57
Operator : JC/SP
Sample : K3156-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 60 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HD-02-060519

Quant Time: Jun 10 02:16:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

