

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060819\
 Data File : VN056147.D
 Acq On : 8 Jun 2019 12:27
 Operator : JC/SP
 Sample : K3225-09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 66 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP19-CHY07-9001

Quant Time: Jun 10 02:31:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	323610	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	496288	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	474699	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	156183	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	172939	53.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.64%	
35) Dibromofluoromethane	7.59	113	151596	49.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.58%	
50) Toluene-d8	10.09	98	635287	54.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.44%	
62) 4-Bromofluorobenzene	12.40	95	200755	50.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.76%	
Target Compounds						
16) Acetone	3.82	43	4765	3.101	ug/l	89
27) cis-1,2-Dichloroethene	6.83	96	2364	0.691	ug/l	93
44) Trichloroethene	8.83	130	4928	1.323	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN060819\
Data File : VN056147.D
Acq On : 8 Jun 2019 12:27
Operator : JC/SP
Sample : K3225-09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 66 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SP19-CHY07-9001

Quant Time: Jun 10 02:31:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

