

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060819\  
 Data File : VN056149.D  
 Acq On : 8 Jun 2019 13:18  
 Operator : JC/SP  
 Sample : K3225-02  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 68 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 SP19-MMW-3414

Quant Time: Jun 10 02:36:40 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N060419W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Jun 04 11:02:09 2019  
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	315159	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	465649	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	450977	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	141504	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	165886	53.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.02%	
35) Dibromofluoromethane	7.59	113	146488	51.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.54%	
50) Toluene-d8	10.09	98	600679	54.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.28%	
62) 4-Bromofluorobenzene	12.40	95	187730	50.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.42%	
Target Compounds						
						Qvalue
16) Acetone	3.82	43	6799	4.543	ug/l #	80
27) cis-1,2-Dichloroethene	6.83	96	4113	1.234	ug/l	83
37) Ethyl Acetate	6.94	43	16507	4.634	ug/l	97
44) Trichloroethene	8.84	130	1891	0.541	ug/l	99
52) Toluene	10.16	92	2937	0.388	ug/l	91
95) Naphthalene	15.14	128	3846	4.376	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060819\  
 Data File : VN056149.D  
 Acq On : 8 Jun 2019 13:18  
 Operator : JC/SP  
 Sample : K3225-02  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 68 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 SP19-MMW-3414

Quant Time: Jun 10 02:36:40 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N060419W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Jun 04 11:02:09 2019  
 Response via : Initial Calibration

