

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060820\
 Data File : VN061687.D
 Acq On : 8 Jun 2020 11:33
 Operator : JC/MD
 Sample : VN0608WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0608WBL01

Quant Time: Jun 09 03:29:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	305756	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	499418	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	459448	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	183614	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.98	65	162079	44.49	ug/l	0.00
Spiked Amount			Recovery	=	88.98%	
35) Dibromofluoromethane	7.54	113	142390	56.20	ug/l	0.00
Spiked Amount			Recovery	=	112.40%	
50) Toluene-d8	10.06	98	581965	49.57	ug/l	0.00
Spiked Amount			Recovery	=	99.14%	
62) 4-Bromofluorobenzene	12.37	95	192815	45.02	ug/l	0.00
Spiked Amount			Recovery	=	90.04%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN060820\
Data File : VN061687.D
Acq On : 8 Jun 2020 11:33
Operator : JC/MD
Sample : VN0608WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0608WBL01

Quant Time: Jun 09 03:29:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

