

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060820\
 Data File : VN061693.D
 Acq On : 8 Jun 2020 14:11
 Operator : JC/MD
 Sample : PB129340TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129340TB

Quant Time: Jun 09 03:55:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	315009	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	534227	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	541513	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	237451	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	175218	46.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.38%	
35) Dibromofluoromethane	7.55	113	139963	51.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.30%	
50) Toluene-d8	10.06	98	651946	51.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.82%	
62) 4-Bromofluorobenzene	12.38	95	233606	50.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.98%	
Target Compounds						
16) Acetone	3.79	43	21949	15.401	ug/l	100
20) Methylene Chloride	4.50	84	19384	4.085	ug/l	90
43) Isopropyl Acetate	8.13	43	61703	9.004	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN060820\
Data File : VN061693.D
Acq On : 8 Jun 2020 14:11
Operator : JC/MD
Sample : PB129340TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129340TB

Quant Time: Jun 09 03:55:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

