

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060820\
 Data File : VN061695.D
 Acq On : 8 Jun 2020 14:57
 Operator : JC/MD
 Sample : L2886-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GEN

Quant Time: Jun 09 04:03:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	320980	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	541532	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	542630	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	246411	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	178717	46.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.46%	
35) Dibromofluoromethane	7.55	113	154076	56.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.18%	
50) Toluene-d8	10.06	98	655556	51.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.00%	
62) 4-Bromofluorobenzene	12.37	95	233393	50.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.50%	
Target Compounds						
16) Acetone	3.78	43	16162	9.366	ug/l	93
20) Methylene Chloride	4.50	84	10666	1.377	ug/l	90
43) Isopropyl Acetate	8.12	43	28097	4.045	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN060820\
Data File : VN061695.D
Acq On : 8 Jun 2020 14:57
Operator : JC/MD
Sample : L2886-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GEN

Quant Time: Jun 09 04:03:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

