

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060822\
 Data File : VN072708.D
 Acq On : 08 Jun 2022 19:05
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jun 09 05:18:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060822W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 05:17:31 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/09/2022
 Supervised By : Mahesh Dadoda 06/09/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	262800	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	445985	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	382796	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	140929	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	195836	27.765	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	55.520%#
35) Dibromofluoromethane	8.016	113	150647	34.137	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	68.280%#
50) Toluene-d8	10.439	98	486449	27.629	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	55.260%#
62) 4-Bromofluorobenzene	12.727	95	184225	30.492	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	60.980%
Target Compounds						
95) Naphthalene	15.504	128	60061m	6.486	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060822\
Data File : VN072708.D
Acq On : 08 Jun 2022 19:05
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Manual Integrations
APPROVED

Reviewed By :John Carlone 06/09/2022
Supervised By :Mahesh Dadoda 06/09/2022

Quant Time: Jun 09 05:18:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060822W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 09 05:17:31 2022
Response via : Initial Calibration

