

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060822\
 Data File : VN072695.D
 Acq On : 08 Jun 2022 10:10
 Operator : JC\MD
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Jun 09 03:21:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N060622W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Jun 06 14:10:15 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	30.000	30.000	0.0	77	0.00
2 M	Dichlorodifluoromethane	50.000	20.824	58.4#	34	0.00
3 M	Chloromethane	50.000	20.074	59.9#	33	0.00
4 M	Vinyl Chloride	50.000	21.104	57.8#	35	0.00
5 M	Bromomethane	50.000	21.728	56.5#	38	0.00
6 M	Chloroethane	50.000	24.026	51.9#	41	0.00
7 M	Trichlorofluoromethane	50.000	20.301	59.4#	33	0.00
8 T	Diethyl Ether	50.000	20.360	59.3#	33	0.00
9	1,1,2-Trichlorotrifluoroeth	50.000	21.570	56.9#	35	0.00
10 M	1,1-Dichloroethene	50.000	19.927	60.1#	33	-0.01
11	Methyl Iodide	50.000	19.734	60.5#	30	0.00
12	Methyl Acetate	50.000	20.723	58.6#	34	0.00
13 M	Acrolein	250.000	99.718	60.1#	34	0.00
14 M	Acrylonitrile	250.000	106.113	57.6#	34	0.00
15 M	Acetone	250.000	104.599	58.2#	34	0.00
16 M	Carbon Disulfide	50.000	18.770	62.5#	32	-0.01
17	Allyl chloride	50.000	20.569	58.9#	33	0.00
18 M	Methylene Chloride	50.000	20.549	58.9#	34	0.00
19 M	trans-1,2-Dichloroethene	50.000	20.051	59.9#	33	0.00
20 T	Diisopropyl ether	50.000	21.923	56.2#	35	0.00
21 M	1,1-Dichloroethane	50.000	21.279	57.4#	34	0.00
22 M	cis-1,2-Dichloroethene	50.000	19.896	60.2#	33	0.00
23 M	tert-Butyl Alcohol	250.000	99.173	60.3#	33	0.00
24 M	Methyl tert-Butyl Ether	50.000	20.555	58.9#	33	0.00
25 M	Chloroform	50.000	21.094	57.8#	34	0.00
26	Cyclohexane	50.000	20.651	58.7#	34	0.00
27 s	1,2-Dichloroethane-d4	30.000	30.359	-1.2	78	0.00
28 I	1,4-Difluorobenzene	30.000	30.000	0.0	78	0.00
29	1,1-Dichloropropene	50.000	20.097	59.8#	33	0.00
30 M	2-Butanone	250.000	102.957	58.8#	33	0.00
31	2,2-Dichloropropane	50.000	19.980	60.0#	33	0.00
32 M	1,1,1-Trichloroethane	50.000	19.981	60.0#	33	0.00
33 M	Carbon Tetrachloride	50.000	19.619	60.8#	32	0.00
34 M	Benzene	50.000	20.285	59.4#	33	0.00
35	Methacrylonitrile	50.000	19.650	60.7#	33	0.00
36 M	1,2-Dichloroethane	50.000	20.410	59.2#	34	0.00
37 M	Trichloroethene	50.000	19.924	60.2#	33	0.00
38	Methylcyclohexane	50.000	20.885	58.2#	35	0.00
39 M	1,2-Dichloropropane	50.000	20.560	58.9#	34	0.00
40	Dibromomethane	50.000	20.642	58.7#	34	0.00
41 M	Bromodichloromethane	50.000	20.050	59.9#	33	0.00
42 M	Vinyl Acetate	250.000	102.364	59.1#	33	0.00
43	Ethyl Acetate	50.000	19.479	61.0#	33	0.00
44	Isopropyl Acetate	50.000	20.715	58.6#	34	0.00
45 T	1,4-Dioxane	1000.000	385.243	61.5#	32	0.00
46	Methyl methacrylate	50.000	19.284	61.4#	33	0.00
47	n-amyl Acetate	50.000	18.721	62.6#	29	0.00
48 M	t-1,3-Dichloropropene	50.000	19.304	61.4#	32	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060822\
 Data File : VN072695.D
 Acq On : 08 Jun 2022 10:10
 Operator : JC\MD
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Jun 09 03:21:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N060622W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Jun 06 14:10:15 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	cis-1,3-Dichloropropene	50.000	19.621	60.8#	32	0.00
50 M	1,1,2-Trichloroethane	50.000	20.676	58.6#	34	0.00
51	Ethyl methacrylate	50.000	19.901	60.2#	33	0.00
52	1,3-Dichloropropane	50.000	20.296	59.4#	34	0.00
53 M	Dibromochloromethane	50.000	19.790	60.4#	32	0.00
54 M	1,2-Dibromoethane	50.000	20.207	59.6#	34	0.00
55 M	2-Chloroethyl vinyl ether	250.000	99.231	60.3#	32	0.00
56 M	Bromoform	50.000	19.311	61.4#	31	0.00
57 I	Chlorobenzene-d5	30.000	30.000	0.0	78	0.00
58 M	4-Methyl-2-Pentanone	250.000	104.493	58.2#	34	0.00
59 M	2-Hexanone	250.000	102.209	59.1#	32	0.00
60 S	4-Bromofluorobenzene	30.000	28.719	4.3	72	0.00
61 M	Tetrachloroethene	50.000	21.341	57.3#	35	0.00
62 M	Toluene	50.000	20.528	58.9#	34	0.00
63 S	Toluene-d8	30.000	29.968	0.1	78	0.00
64 M	Chlorobenzene	50.000	20.279	59.4#	33	0.00
65	1,1,1,2-Tetrachloroethane	50.000	20.352	59.3#	33	0.00
66 M	Ethyl Benzene	50.000	20.339	59.3#	33	0.00
67 M	m/p-Xylenes	100.000	40.559	59.4#	33	0.00
68 M	o-Xylene	50.000	20.100	59.8#	33	0.00
69 M	Styrene	50.000	19.389	61.2#	31	0.00
70	Isopropylbenzene	50.000	20.505	59.0#	33	0.00
71 M	1,1,2,2-Tetrachloroethane	50.000	20.568	58.9#	32	0.00
72	1,2,3-Trichloropropane	50.000	20.579	58.8#	34	0.00
73	Bromobenzene	50.000	20.006	60.0#	32	0.00
74	n-propylbenzene	50.000	20.011	60.0#	32	0.00
75	2-Chlorotoluene	50.000	20.344	59.3#	33	0.00
76	1,3,5-Trimethylbenzene	50.000	20.507	59.0#	33	0.00
77	t-1,4-Dichloro-2-butene	50.000	18.862	62.3#	29	0.00
78	4-Chlorotoluene	50.000	19.437	61.1#	31	0.00
79	tert-butylbenzene	50.000	19.721	60.6#	32	0.00
80	1,2,4-Trimethylbenzene	50.000	19.985	60.0#	32	0.00
81	sec-Butylbenzene	50.000	19.985	60.0#	32	0.00
82	p-Isopropyltoluene	50.000	19.990	60.0#	31	0.00
83 M	1,3-Dichlorobenzene	50.000	19.535	60.9#	31	0.00
84 M	1,4-Dichlorobenzene	50.000	19.774	60.5#	32	0.00
85	n-Butylbenzene	50.000	19.254	61.5#	30	0.00
86 T	Hexachloroethane	50.000	19.430	61.1#	30	0.00
87 M	1,2-Dichlorobenzene	50.000	19.621	60.8#	31	0.00
88	1,2-Dibromo-3-Chloropropane	50.000	18.884	62.2#	30	0.00
89	1,2,4-Trichlorobenzene	50.000	18.478	63.0#	29	0.00
90	Hexachlorobutadiene	50.000	19.854	60.3#	31	0.00
91 M	Naphthalene	50.000	17.105	65.8#	27	0.00
92	1,2,3-Trichlorobenzene	50.000	18.560	62.9#	29	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0