

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060822\
 Data File : VN072719.D
 Acq On : 09 Jun 2022 00:46
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 09 05:21:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060822W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 05:17:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	50.000	46.918	6.2	97	0.00
3 P	Chloromethane	50.000	45.742	8.5	92	0.00
4 C	Vinyl Chloride	50.000	45.965	8.1#	97	0.00
5 T	Bromomethane	50.000	51.493	-3.0	96	0.00
6 T	Chloroethane	50.000	50.058	-0.1	99	0.00
7 T	Trichlorofluoromethane	50.000	46.808	6.4	99	0.00
8 T	Diethyl Ether	50.000	47.291	5.4	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	45.518	9.0	97	0.00
10 T	Methyl Iodide	50.000	49.669	0.7	97	0.00
11 T	Tert butyl alcohol	250.000	235.177	5.9	97	0.00
12 CM	1,1-Dichloroethene	50.000	46.195	7.6#	98	0.00
13 T	Acrolein	250.000	226.788	9.3	94	0.00
14 T	Allyl chloride	50.000	45.400	9.2	95	0.00
15 T	Acrylonitrile	250.000	240.121	4.0	96	0.00
16 T	Acetone	250.000	237.996	4.8	99	0.00
17 T	Carbon Disulfide	50.000	46.102	7.8	94	0.00
18 T	Methyl Acetate	50.000	46.416	7.2	97	0.00
19 T	Methyl tert-butyl Ether	50.000	48.601	2.8	100	0.00
20 T	Methylene Chloride	50.000	48.989	2.0	98	0.00
21 T	trans-1,2-Dichloroethene	50.000	46.307	7.4	97	0.00
22 T	Diisopropyl ether	50.000	48.582	2.8	100	0.00
23 T	Vinyl Acetate	250.000	243.694	2.5	99	0.00
24 P	1,1-Dichloroethane	50.000	48.437	3.1	99	0.00
25 T	2-Butanone	250.000	237.154	5.1	98	0.00
26 T	2,2-Dichloropropane	50.000	36.638	26.7#	79	0.00
27 T	cis-1,2-Dichloroethene	50.000	47.146	5.7	96	0.00
28 T	Bromochloromethane	50.000	50.078	-0.2	101	0.00
29 T	Tetrahydrofuran	250.000	238.910	4.4	99	0.00
30 C	Chloroform	50.000	48.521	3.0#	100	0.00
31 T	Cyclohexane	50.000	48.684	2.6	98	0.00
32 T	1,1,1-Trichloroethane	50.000	48.680	2.6	99	0.00
33 S	1,2-Dichloroethane-d4	50.000	50.786	-1.6	106	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	104	0.00
35 S	Dibromofluoromethane	50.000	49.917	0.2	105	0.00
36 T	1,1-Dichloropropene	50.000	47.163	5.7	99	0.00
37 T	Ethyl Acetate	50.000	47.311	5.4	98	0.00
38 T	Carbon Tetrachloride	50.000	48.716	2.6	101	0.00
39 T	Methylcyclohexane	50.000	45.480	9.0	96	0.00
40 TM	Benzene	50.000	47.093	5.8	99	0.00
41 T	Methacrylonitrile	50.000	45.910	8.2	96	0.00
42 TM	1,2-Dichloroethane	50.000	47.160	5.7	98	0.00
43 T	Isopropyl Acetate	50.000	47.318	5.4	99	0.00
44 TM	Trichloroethene	50.000	47.564	4.9	98	0.00
45 C	1,2-Dichloropropane	50.000	47.943	4.1#	99	0.00
46 T	Dibromomethane	50.000	47.884	4.2	101	0.00
47 T	Bromodichloromethane	50.000	49.219	1.6	100	0.00
48 T	Methyl methacrylate	50.000	44.203	11.6	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060822\
 Data File : VN072719.D
 Acq On : 09 Jun 2022 00:46
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 09 05:21:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060822W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 05:17:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	948.941	5.1	92	0.00
50 S	Toluene-d8	50.000	48.353	3.3	102	0.00
51 T	4-Methyl-2-Pentanone	250.000	239.431	4.2	97	0.00
52 CM	Toluene	50.000	47.156	5.7#	96	0.00
53 T	t-1,3-Dichloropropene	50.000	45.777	8.4	91	0.00
54 T	cis-1,3-Dichloropropene	50.000	46.296	7.4	95	0.00
55 T	1,1,2-Trichloroethane	50.000	46.276	7.4	97	0.00
56 T	Ethyl methacrylate	50.000	47.070	5.9	95	0.00
57 T	1,3-Dichloropropane	50.000	46.823	6.4	97	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	244.403	2.2	94	0.00
59 T	2-Hexanone	250.000	242.399	3.0	94	0.00
60 T	Dibromochloromethane	50.000	47.893	4.2	95	0.00
61 T	1,2-Dibromoethane	50.000	47.164	5.7	96	0.00
62 S	4-Bromofluorobenzene	50.000	48.874	2.3	100	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	98	0.00
64 T	Tetrachloroethene	50.000	46.966	6.1	98	0.00
65 PM	Chlorobenzene	50.000	48.425	3.2	95	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	49.039	1.9	96	0.00
67 C	Ethyl Benzene	50.000	48.892	2.2#	95	0.00
68 T	m/p-Xylenes	100.000	97.397	2.6	95	0.00
69 T	o-Xylene	50.000	48.977	2.0	94	0.00
70 T	Styrene	50.000	50.752	-1.5	95	0.00
71 P	Bromoform	50.000	49.699	0.6	91	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	99	0.00
73 T	Isopropylbenzene	50.000	45.087	9.8	94	0.00
74 T	N-ethyl acetate	50.000	49.106	1.8	94	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	45.830	8.3	93	0.00
76 T	1,2,3-Trichloropropane	50.000	44.497	11.0	91	0.00
77 T	Bromobenzene	50.000	46.014	8.0	95	0.00
78 T	n-propylbenzene	50.000	46.647	6.7	93	0.00
79 T	2-Chlorotoluene	50.000	46.559	6.9	95	0.00
80 T	1,3,5-Trimethylbenzene	50.000	46.489	7.0	94	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	41.152	17.7	81	0.00
82 T	4-Chlorotoluene	50.000	46.702	6.6	94	0.00
83 T	tert-Butylbenzene	50.000	46.693	6.6	95	0.00
84 T	1,2,4-Trimethylbenzene	50.000	47.590	4.8	95	0.00
85 T	sec-Butylbenzene	50.000	47.416	5.2	94	0.00
86 T	p-Isopropyltoluene	50.000	47.980	4.0	94	0.00
87 T	1,3-Dichlorobenzene	50.000	46.784	6.4	97	0.00
88 T	1,4-Dichlorobenzene	50.000	46.186	7.6	94	0.00
89 T	n-Butylbenzene	50.000	47.775	4.5	93	0.00
90 T	Hexachloroethane	50.000	46.592	6.8	90	0.00
91 T	1,2-Dichlorobenzene	50.000	47.093	5.8	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	44.273	11.5	93	0.00
93 T	1,2,4-Trichlorobenzene	50.000	47.569	4.9	92	0.00
94 T	Hexachlorobutadiene	50.000	47.093	5.8	93	0.00
95 T	Naphthalene	50.000	50.431	-0.9	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060822\
Data File : VN072719.D
Acq On : 09 Jun 2022 00:46
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jun 09 05:21:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060822W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 09 05:17:31 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	49.389	1.2	94	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6