

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060823\
 Data File : VN078117.D
 Acq On : 08 Jun 2023 09:47
 Operator : JC\MD
 Sample : VN0608MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0608MBL01

Quant Time: Jun 09 01:21:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	348410	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	626439	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	552987	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	161210	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	237590	51.146	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.300%
35) Dibromofluoromethane	8.171	113	209974	53.012	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.020%
50) Toluene-d8	10.571	98	773988	50.688	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.380%
62) 4-Bromofluorobenzene	12.853	95	217564	45.832	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060823\
Data File : VN078117.D
Acq On : 08 Jun 2023 09:47
Operator : JC\MD
Sample : VN0608MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0608MBL01

Quant Time: Jun 09 01:21:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

