

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060823\
 Data File : VN078123.D
 Acq On : 08 Jun 2023 12:37
 Operator : JC\MD
 Sample : 03050-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-23-0A

Quant Time: Jun 09 01:25:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	373630	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	705737	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	644409	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	187276	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	260307	52.253	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 104.500%		
35) Dibromofluoromethane	8.171	113	226306	50.715	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.440%		
50) Toluene-d8	10.571	98	878003	51.039	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 102.080%		
62) 4-Bromofluorobenzene	12.853	95	252945	47.298	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 94.600%		
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	41378	28.857	ug/l	89
20) Methylene Chloride	5.295	84	15552	3.282	ug/l #	85
37) Ethyl Acetate	7.577	43	58643	11.728	ug/l	97
43) Isopropyl Acetate	8.694	43	332871	34.810	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060823\
Data File : VN078123.D
Acq On : 08 Jun 2023 12:37
Operator : JC\MD
Sample : 03050-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-23-0A

Quant Time: Jun 09 01:25:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

