

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060823\
 Data File : VN078124.D
 Acq On : 08 Jun 2023 13:00
 Operator : JC\MD
 Sample : 03092-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1

Quant Time: Jun 09 01:25:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	366100	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	684591	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	636530	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	191639	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	258739	53.007	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.020%
35) Dibromofluoromethane	8.177	113	223862	51.717	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.440%
50) Toluene-d8	10.571	98	873937	52.372	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.740%
62) 4-Bromofluorobenzene	12.853	95	251613	48.503	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.000%
Target Compounds						
						Qvalue
16) Acetone	4.447	43	41957	29.863	ug/l	97
20) Methylene Chloride	5.294	84	19320	4.161	ug/l	92
37) Ethyl Acetate	7.571	43	59418	12.250	ug/l	97
43) Isopropyl Acetate	8.694	43	341627	36.829	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060823\
Data File : VN078124.D
Acq On : 08 Jun 2023 13:00
Operator : JC\MD
Sample : 03092-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1

Quant Time: Jun 09 01:25:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

