

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060823\
 Data File : VN078127.D
 Acq On : 08 Jun 2023 14:12
 Operator : JC\MD
 Sample : PB153290TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB153290TB

Quant Time: Jun 09 01:26:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	354496	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	654731	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	612162	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	186350	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	243058	51.424	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.840%
35) Dibromofluoromethane	8.171	113	210157	50.765	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.540%
50) Toluene-d8	10.571	98	824917	51.689	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.380%
62) 4-Bromofluorobenzene	12.853	95	245154	49.413	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.820%
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	29883	21.965	ug/l	99
20) Methylene Chloride	5.289	84	9557	2.126	ug/l	93
37) Ethyl Acetate	7.571	43	63315	13.649	ug/l	96
43) Isopropyl Acetate	8.694	43	360947	40.686	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060823\
Data File : VN078127.D
Acq On : 08 Jun 2023 14:12
Operator : JC\MD
Sample : PB153290TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB153290TB

Quant Time: Jun 09 01:26:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

