

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060823\
 Data File : VN078130.D
 Acq On : 08 Jun 2023 15:36
 Operator : JC\MD
 Sample : 03072-04 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 30998

Quant Time: Jun 09 01:27:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	350931	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	653748	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	582854	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	183571	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	244930	52.347	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.700%
35) Dibromofluoromethane	8.177	113	213143	51.564	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.120%
50) Toluene-d8	10.571	98	807395	50.667	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.340%
62) 4-Bromofluorobenzene	12.853	95	238952	48.235	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060823\
Data File : VN078130.D
Acq On : 08 Jun 2023 15:36
Operator : JC\MD
Sample : 03072-04 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
30998

Quant Time: Jun 09 01:27:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

