

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060920\
 Data File : VN061716.D
 Acq On : 9 Jun 2020 10:58
 Operator : JC/MD
 Sample : VN0609WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0609WBL01

Quant Time: Jun 10 03:27:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	271361	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	445829	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	422856	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	167048	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	149798	46.33	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.66%	
35) Dibromofluoromethane	7.54	113	130610	57.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	115.50%	
50) Toluene-d8	10.06	98	526807	50.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.54%	
62) 4-Bromofluorobenzene	12.38	95	173419	45.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.70%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN060920\
Data File : VN061716.D
Acq On : 9 Jun 2020 10:58
Operator : JC/MD
Sample : VN0609WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0609WBL01

Quant Time: Jun 10 03:27:13 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

