

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060920\
 Data File : VN061721.D
 Acq On : 9 Jun 2020 13:17
 Operator : JC/MD
 Sample : PB129410ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129410ZHE#02

Quant Time: Jun 10 03:46:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	280325	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	477047	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	489118	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	210205	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	161246	48.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.56%	
35) Dibromofluoromethane	7.55	113	129778	53.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.26%	
50) Toluene-d8	10.06	98	591450	52.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.48%	
62) 4-Bromofluorobenzene	12.38	95	206390	50.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.88%	
Target Compounds						
16) Acetone	3.78	43	13602	8.795	ug/l	96
20) Methylene Chloride	4.51	84	8848	1.217	ug/l	97
43) Isopropyl Acetate	8.13	43	64312	10.509	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060920\
Data File : VN061721.D
Acq On : 9 Jun 2020 13:17
Operator : JC/MD
Sample : PB129410ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129410ZHE#02

Quant Time: Jun 10 03:46:15 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

