

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN060920\
 Data File : VN061723.D
 Acq On : 9 Jun 2020 14:03
 Operator : JC/MD
 Sample : L2946-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CUTTINGS

Quant Time: Jun 10 03:50:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	301516	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	515929	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	523393	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	228320	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	172050	47.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.78%	
35) Dibromofluoromethane	7.55	113	150089	57.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.70%	
50) Toluene-d8	10.06	98	634284	52.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.60%	
62) 4-Bromofluorobenzene	12.38	95	222667	50.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.64%	
Target Compounds						
16) Acetone	3.78	43	20954	15.344	ug/l	95
20) Methylene Chloride	4.51	84	13134	2.365	ug/l	91
43) Isopropyl Acetate	8.13	43	51534	7.787	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN060920\
Data File : VN061723.D
Acq On : 9 Jun 2020 14:03
Operator : JC/MD
Sample : L2946-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CUTTINGS

Quant Time: Jun 10 03:50:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

