

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060922\
 Data File : VN072726.D
 Acq On : 09 Jun 2022 13:25
 Operator : JC\MD
 Sample : N3215-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 N48964

Quant Time: Jun 09 15:49:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060822W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 05:17:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	171013	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	324481	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	292334	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	108837	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	265433	57.830	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	115.660%
35) Dibromofluoromethane	8.016	113	185724	57.846	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	115.700%
50) Toluene-d8	10.439	98	639110	49.892	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.780%
62) 4-Bromofluorobenzene	12.727	95	222029	50.510	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.020%
Target Compounds						
						Qvalue
16) Acetone	4.298	43	10591	4.816	ug/l	# 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060922\
Data File : VN072726.D
Acq On : 09 Jun 2022 13:25
Operator : JC\MD
Sample : N3215-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
N48964

Quant Time: Jun 09 15:49:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060822W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 09 05:17:31 2022
Response via : Initial Calibration

