

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060922\
 Data File : VN072727.D
 Acq On : 09 Jun 2022 13:50
 Operator : JC\MD
 Sample : N3170-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20220409-FIELD-BLANK-2

Quant Time: Jun 09 15:50:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060822W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 05:17:31 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	176139	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	335952	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	297138	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	106832	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	269046	56.911	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	113.820%
35) Dibromofluoromethane	8.016	113	191627	57.646	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	115.300%
50) Toluene-d8	10.439	98	641851	48.395	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.780%
62) 4-Bromofluorobenzene	12.727	95	215586	47.370	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.740%
Target Compounds						
						Qvalue
16) Acetone	4.293	43	14199	6.268	ug/l	90
20) Methylene Chloride	5.081	84	15240	4.521	ug/l	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060922\
Data File : VN072727.D
Acq On : 09 Jun 2022 13:50
Operator : JC\MD
Sample : N3170-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20220409-FIELD-BLANK-2

Quant Time: Jun 09 15:50:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060822W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 09 05:17:31 2022
Response via : Initial Calibration

