

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060922\
 Data File : VN072738.D
 Acq On : 09 Jun 2022 18:18
 Operator : JC\MD
 Sample : N3175-21RE 20X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1713RE

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/10/2022
 Supervised By : Mahesh Dadoda 06/10/2022

Quant Time: Jun 09 18:37:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060822W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 05:17:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	304712	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	577208	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	521497	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	206452	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	412885	50.486	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.980%	
35) Dibromofluoromethane	8.016	113	297089	52.017	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 104.040%	
50) Toluene-d8	10.439	98	990401	43.463	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 86.920%	
62) 4-Bromofluorobenzene	12.727	95	378984	48.467	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 96.940%	
Target Compounds						
11) Tert butyl alcohol	5.381	59	11036m	5.657	ug/l	Qvalue
16) Acetone	4.287	43	32263	8.233	ug/l	# 82
18) Methyl Acetate	4.857	43	18874	1.851	ug/l	# 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060922\
Data File : VN072738.D
Acq On : 09 Jun 2022 18:18
Operator : JC\MD
Sample : N3175-21RE 20X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1713RE

Quant Time: Jun 09 18:37:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060822W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 09 05:17:31 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 06/10/2022
Supervised By :Mahesh Dadoda 06/10/2022

