

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060922\
 Data File : VN072743.D
 Acq On : 09 Jun 2022 20:20
 Operator : JC\MD
 Sample : N3280-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 33RD-AVE-CASING-10

Quant Time: Jun 10 01:47:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060822W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 05:17:31 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	160441	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	311956	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	273416	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	98253	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	243120	56.459	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	112.920%
35) Dibromofluoromethane	8.022	113	173765	56.294	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	112.580%
50) Toluene-d8	10.439	98	571833	46.432	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	92.860%
62) 4-Bromofluorobenzene	12.727	95	196882	46.588	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.180%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	14516	7.035	ug/l	# 79
19) Methyl tert-butyl Ether	5.610	73	4394	0.420	ug/l	# 89
25) 2-Butanone	7.339	43	5612	1.691	ug/l	94
40) Benzene	8.463	78	4336	0.333	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060922\
Data File : VN072743.D
Acq On : 09 Jun 2022 20:20
Operator : JC\MD
Sample : N3280-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
33RD-AVE-CASING-10

Quant Time: Jun 10 01:47:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060822W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 09 05:17:31 2022
Response via : Initial Calibration

