

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060922\  
 Data File : VN072725.D  
 Acq On : 09 Jun 2022 13:01  
 Operator : JC\MD  
 Sample : VN0609WBSD01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0609WBSD01

Manual Integrations  
 APPROVED

Reviewed By : John Carlone 06/10/2022  
 Supervised By : Mahesh Dadoda 06/10/2022

Quant Time: Jun 09 16:36:02 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060822W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Jun 09 05:17:31 2022  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response            | Conc    | Units | Dev(Min) |
|------------------------------|----------------|------|---------------------|---------|-------|----------|
| -----                        |                |      |                     |         |       |          |
| Internal Standards           |                |      |                     |         |       |          |
| 1) Pentafluorobenzene        | 8.081          | 168  | 148466              | 50.000  | ug/l  | # 0.00   |
| 34) 1,4-Difluorobenzene      | 8.963          | 114  | 274397              | 50.000  | ug/l  | 0.00     |
| 63) Chlorobenzene-d5         | 11.739         | 117  | 295243              | 50.000  | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 13.668         | 152  | 134056              | 50.000  | ug/l  | 0.00     |
| System Monitoring Compounds  |                |      |                     |         |       |          |
| 33) 1,2-Dichloroethane-d4    | 8.428          | 65   | 209486              | 52.572  | ug/l  | 0.00     |
| Spiked Amount 50.000         | Range 61 - 141 |      | Recovery = 105.140% |         |       |          |
| 35) Dibromofluoromethane     | 8.016          | 113  | 139347              | 51.323  | ug/l  | 0.00     |
| Spiked Amount 50.000         | Range 69 - 133 |      | Recovery = 102.640% |         |       |          |
| 50) Toluene-d8               | 10.439         | 98   | 548043              | 50.591  | ug/l  | 0.00     |
| Spiked Amount 50.000         | Range 65 - 126 |      | Recovery = 101.180% |         |       |          |
| 62) 4-Bromofluorobenzene     | 12.727         | 95   | 192823              | 51.873  | ug/l  | 0.00     |
| Spiked Amount 50.000         | Range 58 - 135 |      | Recovery = 103.740% |         |       |          |
| Target Compounds             |                |      |                     |         |       |          |
|                              |                |      |                     |         |       | Qvalue   |
| 2) Dichlorodifluoromethane   | 2.069          | 85   | 46858               | 19.456  | ug/l  | 100      |
| 3) Chloromethane             | 2.293          | 50   | 51863               | 19.670  | ug/l  | 96       |
| 4) Vinyl Chloride            | 2.434          | 62   | 48566               | 18.415  | ug/l  | 99       |
| 5) Bromomethane              | 2.828          | 94   | 28668               | 22.592  | ug/l  | 98       |
| 6) Chloroethane              | 3.004          | 64   | 37921               | 22.273  | ug/l  | 90       |
| 7) Trichlorofluoromethane    | 3.363          | 101  | 82510               | 19.657  | ug/l  | 95       |
| 8) Diethyl Ether             | 3.816          | 74   | 35115               | 20.575  | ug/l  | 80       |
| 9) 1,1,2-Trichlorotrifluo... | 4.199          | 101  | 45864               | 19.953  | ug/l  | 89       |
| 10) Methyl Iodide            | 4.416          | 142  | 40524               | 18.908  | ug/l  | # 85     |
| 11) Tert butyl alcohol       | 5.363          | 59   | 96116               | 101.116 | ug/l  | 100      |
| 12) 1,1-Dichloroethene       | 4.175          | 96   | 41104               | 19.187  | ug/l  | 89       |
| 13) Acrolein                 | 4.034          | 56   | 41192               | 104.926 | ug/l  | 98       |
| 14) Allyl chloride           | 4.834          | 41   | 98491               | 19.914  | ug/l  | # 90     |
| 15) Acrylonitrile            | 5.545          | 53   | 209817              | 104.180 | ug/l  | 99       |
| 16) Acetone                  | 4.281          | 43   | 191200              | 100.138 | ug/l  | 92       |
| 17) Carbon Disulfide         | 4.522          | 76   | 93160               | 19.017  | ug/l  | 99       |
| 18) Methyl Acetate           | 4.851          | 43   | 104661              | 21.067  | ug/l  | # 89     |
| 19) Methyl tert-butyl Ether  | 5.610          | 73   | 206817              | 21.344  | ug/l  | 98       |
| 20) Methylene Chloride       | 5.093          | 84   | 58117               | 21.202  | ug/l  | 87       |
| 21) trans-1,2-Dichloroethene | 5.587          | 96   | 45160               | 19.144  | ug/l  | 89       |
| 22) Diisopropyl ether        | 6.487          | 45   | 212359              | 21.185  | ug/l  | 94       |
| 23) Vinyl Acetate            | 6.428          | 43   | 977040              | 106.460 | ug/l  | # 92     |
| 24) 1,1-Dichloroethane       | 6.381          | 63   | 108150              | 20.777  | ug/l  | 97       |
| 25) 2-Butanone               | 7.328          | 43   | 313129              | 101.947 | ug/l  | # 87     |
| 26) 2,2-Dichloropropane      | 7.322          | 77   | 94848               | 20.152  | ug/l  | 95       |
| 27) cis-1,2-Dichloroethene   | 7.322          | 96   | 61874               | 20.622  | ug/l  | 87       |
| 28) Bromochloromethane       | 7.651          | 49   | 53095               | 21.596  | ug/l  | # 78     |
| 29) Tetrahydrofuran          | 7.675          | 42   | 192799              | 101.584 | ug/l  | 90       |
| 30) Chloroform               | 7.810          | 83   | 112422              | 20.718  | ug/l  | 99       |
| 31) Cyclohexane              | 8.086          | 56   | 87548               | 20.330  | ug/l  | # 85     |
| 32) 1,1,1-Trichloroethane    | 8.010          | 97   | 93923               | 19.587  | ug/l  | 96       |
| 36) 1,1-Dichloropropene      | 8.216          | 75   | 72643               | 19.275  | ug/l  | 95       |
| 37) Ethyl Acetate            | 7.404          | 43   | 120273              | 20.541  | ug/l  | 95       |
| 38) Carbon Tetrachloride     | 8.204          | 117  | 80220               | 19.918  | ug/l  | 98       |
| 39) Methylcyclohexane        | 9.457          | 83   | 81736               | 18.689  | ug/l  | 91       |
| 40) Benzene                  | 8.457          | 78   | 231851              | 20.273  | ug/l  | 99       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060922\  
 Data File : VN072725.D  
 Acq On : 09 Jun 2022 13:01  
 Operator : JC\MD  
 Sample : VN0609WBSD01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0609WBSD01

Manual Integrations  
 APPROVED

Reviewed By :John Carlone 06/10/2022  
 Supervised By :Mahesh Dadoda 06/10/2022

Quant Time: Jun 09 16:36:02 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060822W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Jun 09 05:17:31 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 41) Methacrylonitrile         | 7.628  | 41   | 57620    | 29.194  | ug/l   | 90       |
| 42) 1,2-Dichloroethane        | 8.522  | 62   | 102226   | 21.559  | ug/l   | 96       |
| 43) Isopropyl Acetate         | 8.557  | 43   | 194103   | 21.286  | ug/l # | 93       |
| 44) Trichloroethene           | 9.210  | 130  | 53629    | 20.142  | ug/l   | 89       |
| 45) 1,2-Dichloropropane       | 9.486  | 63   | 64660    | 20.923  | ug/l   | 97       |
| 46) Dibromomethane            | 9.575  | 93   | 44625    | 20.656  | ug/l # | 85       |
| 47) Bromodichloromethane      | 9.757  | 83   | 93077    | 20.958  | ug/l   | 97       |
| 48) Methyl methacrylate       | 9.557  | 41   | 85105    | 19.726  | ug/l # | 85       |
| 49) 1,4-Dioxane               | 9.563  | 88   | 34675    | 439.688 | ug/l # | 89       |
| 51) 4-Methyl-2-Pentanone      | 10.328 | 43   | 632853   | 106.568 | ug/l   | 91       |
| 52) Toluene                   | 10.504 | 92   | 146175   | 20.471  | ug/l   | 99       |
| 53) t-1,3-Dichloropropene     | 10.716 | 75   | 104264   | 21.134  | ug/l   | 99       |
| 54) cis-1,3-Dichloropropene   | 10.186 | 75   | 109535   | 21.488  | ug/l # | 84       |
| 55) 1,1,2-Trichloroethane     | 10.898 | 97   | 64504    | 20.933  | ug/l   | 95       |
| 56) Ethyl methacrylate        | 10.757 | 69   | 117985   | 21.649  | ug/l # | 81       |
| 57) 1,3-Dichloropropane       | 11.039 | 76   | 116795   | 21.226  | ug/l   | 99       |
| 58) 2-Chloroethyl Vinyl ether | 10.039 | 63   | 238496   | 109.803 | ug/l   | 92       |
| 59) 2-Hexanone                | 11.080 | 43   | 477482   | 109.344 | ug/l   | 90       |
| 60) Dibromochloromethane      | 11.233 | 129  | 69843    | 21.838  | ug/l   | 98       |
| 61) 1,2-Dibromoethane         | 11.345 | 107  | 68450    | 21.675  | ug/l   | 98       |
| 64) Tetrachloroethene         | 10.975 | 164  | 47385    | 19.063  | ug/l   | 91       |
| 65) Chlorobenzene             | 11.769 | 112  | 159975   | 20.334  | ug/l   | 99       |
| 66) 1,1,1,2-Tetrachloroethane | 11.839 | 131  | 63140    | 20.175  | ug/l   | 95       |
| 67) Ethyl Benzene             | 11.839 | 91   | 293200   | 19.820  | ug/l   | 97       |
| 68) m/p-Xylenes               | 11.951 | 106  | 211459   | 40.224  | ug/l   | 89       |
| 69) o-Xylene                  | 12.280 | 106  | 108959   | 20.618  | ug/l   | 92       |
| 70) Styrene                   | 12.292 | 104  | 172074   | 20.718  | ug/l   | 95       |
| 71) Bromoform                 | 12.457 | 173  | 46706    | 21.218  | ug/l # | 97       |
| 73) Isopropylbenzene          | 12.574 | 105  | 281484   | 19.575  | ug/l   | 97       |
| 74) N-amyl acetate            | 12.386 | 43   | 143102   | 21.412  | ug/l # | 88       |
| 75) 1,1,2,2-Tetrachloroethane | 12.821 | 83   | 102382   | 20.511  | ug/l   | 100      |
| 76) 1,2,3-Trichloropropane    | 12.874 | 75   | 102654m  | 23.730  | ug/l   |          |
| 77) Bromobenzene              | 12.857 | 156  | 57515    | 20.290  | ug/l   | 68       |
| 78) n-propylbenzene           | 12.916 | 91   | 317860   | 20.246  | ug/l   | 94       |
| 79) 2-Chlorotoluene           | 13.004 | 91   | 195985   | 19.899  | ug/l   | 94       |
| 80) 1,3,5-Trimethylbenzene    | 13.057 | 105  | 232925   | 20.443  | ug/l   | 96       |
| 81) trans-1,4-Dichloro-2-b... | 12.621 | 75   | 32624    | 19.723  | ug/l # | 85       |
| 82) 4-Chlorotoluene           | 13.104 | 91   | 185879   | 20.434  | ug/l   | 94       |
| 83) tert-Butylbenzene         | 13.321 | 119  | 198274   | 19.813  | ug/l   | 88       |
| 84) 1,2,4-Trimethylbenzene    | 13.363 | 105  | 225961   | 20.727  | ug/l   | 94       |
| 85) sec-Butylbenzene          | 13.498 | 105  | 275395   | 20.395  | ug/l   | 96       |
| 86) p-Isopropyltoluene        | 13.610 | 119  | 215953   | 20.331  | ug/l   | 95       |
| 87) 1,3-Dichlorobenzene       | 13.616 | 146  | 98980    | 20.558  | ug/l   | 97       |
| 88) 1,4-Dichlorobenzene       | 13.692 | 146  | 97747    | 20.723  | ug/l   | 96       |
| 89) n-Butylbenzene            | 13.939 | 91   | 182025   | 20.340  | ug/l   | 96       |
| 90) Hexachloroethane          | 14.204 | 117  | 43566    | 18.950  | ug/l   | 75       |
| 91) 1,2-Dichlorobenzene       | 13.986 | 146  | 98114    | 20.791  | ug/l   | 95       |
| 92) 1,2-Dibromo-3-Chloropr... | 14.598 | 75   | 23619    | 19.589  | ug/l   | 70       |
| 93) 1,2,4-Trichlorobenzene    | 15.262 | 180  | 47751    | 21.034  | ug/l   | 93       |
| 94) Hexachlorobutadiene       | 15.368 | 225  | 21563    | 21.204  | ug/l   | 96       |
| 95) Naphthalene               | 15.504 | 128  | 183161   | 20.795  | ug/l   | 99       |
| 96) 1,2,3-Trichlorobenzene    | 15.698 | 180  | 49714    | 21.991  | ug/l   | 96       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN060922\  
Data File : VN072725.D  
Acq On : 09 Jun 2022 13:01  
Operator : JC\MD  
Sample : VN0609WBSD01  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 6 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01

Manual Integrations  
APPROVED


Reviewed By :John Carlone 06/10/2022  
Supervised By :Mahesh Dadoda 06/10/2022

Quant Time: Jun 09 16:36:02 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060822W.M  
Quant Title : SW846 8260  
QLast Update : Thu Jun 09 05:17:31 2022  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|


-----  
(#) = qualifier out of range (m) = manual integration (+) = signals summed





#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 8.081 min Scan# 1036  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01



Tgt Ion:168 Resp: 148460  
Ion Ratio Lower Upper  
168 100  
99 71.9 42.3 63.5




Manual Integrations  
APPROVED


Reviewed By :John Carlone 06/10/2022  
Supervised By :Mahesh Dadoda 06/10/2022



#2  
Dichlorodifluoromethane  
Concen: 19.456 ug/l  
RT: 2.069 min Scan# 14  
Delta R.T. 0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01


Tgt Ion: 85 Resp: 46858  
Ion Ratio Lower Upper  
85 100  
87 31.5 15.7 47.1





#3  
Chloromethane  
Concen: 19.670 ug/l  
RT: 2.293 min Scan# 51  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01




Tgt Ion: 50 Resp: 5186  
Ion Ratio Lower Upper  
50 100  
52 34.5 25.7 38.5

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/10/2022  
Supervised By :Mahesh Dadoda 06/10/2022




#4  
Vinyl Chloride  
Concen: 18.415 ug/l  
RT: 2.434 min Scan# 76  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01




Tgt Ion: 62 Resp: 48566  
Ion Ratio Lower Upper  
62 100  
64 30.5 25.0 37.4





#5  
Bromomethane  
Concen: 22.592 ug/l  
RT: 2.828 min Scan# 141  
Delta R.T. 0.012 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01



Tgt Ion: 94 Resp: 2866  
Ion Ratio Lower Upper  
94 100  
96 96.6 76.0 114.0




Manual Integrations  
APPROVED


Reviewed By :John Carlone 06/10/2022  
Supervised By :Mahesh Dadoda 06/10/2022



#6  
Chloroethane  
Concen: 22.273 ug/l  
RT: 3.004 min Scan# 173  
Delta R.T. 0.011 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01

Tgt Ion: 64 Resp: 37921  
Ion Ratio Lower Upper  
64 100  
66 27.1 26.2 39.4









#7  
Trichlorofluoromethane  
Concen: 19.657 ug/l  
RT: 3.363 min Scan# 21  
Delta R.T. 0.006 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01

Tgt Ion:101 Resp: 82510  
Ion Ratio Lower Upper  
101 100  
103 60.8 51.5 77.3




Manual Integrations  
APPROVED

Reviewed By :John Carlone 06/10/2022  
Supervised By :Mahesh Dadoda 06/10/2022




#8  
Diethyl Ether  
Concen: 20.575 ug/l  
RT: 3.816 min Scan# 311  
Delta R.T. 0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01

Tgt Ion: 74 Resp: 35115  
Ion Ratio Lower Upper  
74 100  
45 103.7 42.9 128.5











#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 19.953 ug/l  
RT: 4.199 min Scan# 376  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01

Tgt Ion:101 Resp: 45864  
Ion Ratio Lower Upper  
101 100  
85 48.8 34.7 52.1  
151 67.9 63.0 94.4




Manual Integrations  
APPROVED


Reviewed By :John Carlone 06/10/2022  
Supervised By :Mahesh Dadoda 06/10/2022



#10  
Methyl Iodide  
Concen: 18.908 ug/l  
RT: 4.416 min Scan# 413  
Delta R.T. 0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01


Tgt Ion:142 Resp: 40524  
Ion Ratio Lower Upper  
142 100  
127 51.5 31.3 46.9#  
141 13.7 11.2 16.8





#11  
Tert butyl alcohol  
Concen: 101.116 ug/l  
RT: 5.363 min Scan# 511  
Delta R.T. 0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01




Tgt Ion: 59 Resp: 96110  
Ion Ratio Lower Upper  
59 100  
57 10.5 8.3 12.5

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/10/2022  
Supervised By :Mahesh Dadoda 06/10/2022




#12  
1,1-Dichloroethene  
Concen: 19.187 ug/l  
RT: 4.175 min Scan# 372  
Delta R.T. 0.006 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01




Tgt Ion: 96 Resp: 41104  
Ion Ratio Lower Upper  
96 100  
61 178.6 127.6 191.4  
98 66.3 51.2 76.8





#13  
Acrolein  
Concen: 104.926 ug/l  
RT: 4.034 min Scan# 348  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01




Tgt Ion: 56 Resp: 41192  
Ion Ratio Lower Upper  
56 100  
55 70.2 54.8 82.2

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/10/2022  
Supervised By :Mahesh Dadoda 06/10/2022




#14  
Allyl chloride  
Concen: 19.914 ug/l  
RT: 4.834 min Scan# 484  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01




Tgt Ion: 41 Resp: 98491  
Ion Ratio Lower Upper  
41 100  
39 72.0 53.2 79.8  
76 28.5 29.9 44.9#





#15  
Acrylonitrile  
Concen: 104.180 ug/l  
RT: 5.545 min Scan# 605  
Delta R.T. -0.001 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01



Tgt Ion: 53 Resp: 20981  
Ion Ratio Lower Upper  
53 100  
52 83.7 66.4 99.6  
51 36.4 28.5 42.7




Manual Integrations  
APPROVED


Reviewed By :John Carlone 06/10/2022  
Supervised By :Mahesh Dadoda 06/10/2022



#16  
Acetone  
Concen: 100.138 ug/l  
RT: 4.281 min Scan# 390  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01


Tgt Ion: 43 Resp: 191200  
Ion Ratio Lower Upper  
43 100  
58 28.2 26.0 39.0





#17  
Carbon Disulfide  
Concen: 19.017 ug/l  
RT: 4.522 min Scan# 431  
Delta R.T. 0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01




Tgt Ion: 76 Resp: 93160  
Ion Ratio Lower Upper  
76 100  
78 8.7 7.2 10.8

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/10/2022  
Supervised By :Mahesh Dadoda 06/10/2022




#18  
Methyl Acetate  
Concen: 21.067 ug/l  
RT: 4.851 min Scan# 487  
Delta R.T. 0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01




Tgt Ion: 43 Resp: 104661  
Ion Ratio Lower Upper  
43 100  
74 21.3 21.6 32.4#





#19  
Methyl tert-butyl Ether  
Concen: 21.344 ug/l  
RT: 5.610 min Scan# 616  
Delta R.T. 0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01



Tgt Ion: 73 Resp: 20681  
Ion Ratio Lower Upper  
73 100  
57 21.2 17.6 26.4




Manual Integrations  
APPROVED


Reviewed By :John Carlone 06/10/2022  
Supervised By :Mahesh Dadoda 06/10/2022



#20  
Methylene Chloride  
Concen: 21.202 ug/l  
RT: 5.093 min Scan# 528  
Delta R.T. 0.006 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01


Tgt Ion: 84 Resp: 58117  
Ion Ratio Lower Upper  
84 100  
49 130.1 88.3 132.5  
51 41.5 28.2 42.2  
86 59.3 50.6 76.0





#21  
trans-1,2-Dichloroethene  
Concen: 19.144 ug/l  
RT: 5.587 min Scan# 612  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01




Tgt Ion: 96 Resp: 45160  
Ion Ratio Lower Upper  
96 100  
61 149.6 105.8 158.8  
98 61.0 50.3 75.5

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/10/2022  
Supervised By :Mahesh Dadoda 06/10/2022




#22  
Diisopropyl ether  
Concen: 21.185 ug/l  
RT: 6.487 min Scan# 765  
Delta R.T. -0.006 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01




Tgt Ion: 45 Resp: 212359  
Ion Ratio Lower Upper  
45 100  
43 63.7 46.6 70.0  
87 27.6 24.3 36.5  
59 11.0 9.6 14.4





#23  
 Vinyl Acetate  
 Concen: 106.460 ug/l  
 RT: 6.428 min Scan# 755  
 Delta R.T. -0.000 min  
 Lab File: VN072725.D  
 Acq: 09 Jun 2022 13:01




Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0609WBSD01




Tgt Ion: 43 Resp: 977040  
 Ion Ratio Lower Upper  
 43 100  
 86 9.2 10.0 15.0

Manual Integrations  
 APPROVED



Reviewed By :John Carlone 06/10/2022  
 Supervised By :Mahesh Dadoda 06/10/2022




#24  
 1,1-Dichloroethane  
 Concen: 20.777 ug/l  
 RT: 6.381 min Scan# 747  
 Delta R.T. -0.000 min  
 Lab File: VN072725.D  
 Acq: 09 Jun 2022 13:01



Tgt Ion: 63 Resp: 108150  
 Ion Ratio Lower Upper  
 63 100  
 98 6.1 3.5 10.4  
 100 3.2 2.3 6.8








#25  
2-Butanone  
Concen: 101.947 ug/l  
RT: 7.328 min Scan# 908  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01




Tgt Ion: 43 Resp: 313129  
Ion Ratio Lower Upper  
43 100  
72 22.8 23.7 35.5

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/10/2022  
Supervised By :Mahesh Dadoda 06/10/2022




#26  
2,2-Dichloropropane  
Concen: 20.152 ug/l  
RT: 7.322 min Scan# 907  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01




Tgt Ion: 77 Resp: 94848  
Ion Ratio Lower Upper  
77 100  
97 20.4 11.4 34.2





#27  
 cis-1,2-Dichloroethene  
 Concen: 20.622 ug/l  
 RT: 7.322 min Scan# 906  
 Delta R.T. 0.006 min  
 Lab File: VN072725.D  
 Acq: 09 Jun 2022 13:01




Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0609WBSD01



Tgt Ion: 96 Resp: 61874  
 Ion Ratio Lower Upper  
 96 100  
 61 162.7 0.0 280.6  
 98 63.5 0.0 128.4




Manual Integrations  
 APPROVED


Reviewed By :John Carlone 06/10/2022  
 Supervised By :Mahesh Dadoda 06/10/2022



#28  
 Bromochloromethane  
 Concen: 21.596 ug/l  
 RT: 7.651 min Scan# 963  
 Delta R.T. 0.000 min  
 Lab File: VN072725.D  
 Acq: 09 Jun 2022 13:01

Tgt Ion: 49 Resp: 53095  
 Ion Ratio Lower Upper  
 49 100  
 129 1.8 0.0 4.6  
 130 65.9 69.7 104.5#





#29

Tetrahydrofuran

Concen: 101.584 ug/l

RT: 7.675 min Scan# 90

Delta R.T. -0.006 min

Lab File: VN072725.D


Acq: 09 Jun 2022 13:01

Instrument :

MSVOA\_N

ClientSampleId :

VN0609WBSD01






| Tgt Ion:  | 42   | Resp: | 19279 |
|-----------|------|-------|-------|
| Ion Ratio | 100  | Lower | Upper |
| 42        | 100  |       |       |
| 72        | 42.7 | 39.7  | 59.5  |
| 71        | 39.4 | 36.6  | 55.0  |

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/10/2022

Supervised By :Mahesh Dadoda 06/10/2022



#30

Chloroform

Concen: 20.718 ug/l




RT: 7.810 min Scan# 990


Delta R.T. -0.006 min

Lab File: VN072725.D

Acq: 09 Jun 2022 13:01

| Tgt Ion:  | 83   | Resp: | 112422 |
|-----------|------|-------|--------|
| Ion Ratio | 100  | Lower | Upper  |
| 83        | 100  |       |        |
| 85        | 64.1 | 51.8  | 77.8   |





#31

Cyclohexane

Concen: 20.330 ug/l

RT: 8.086 min Scan# 1037

Delta R.T. -0.006 min

Lab File: VN072725.D

Acq: 09 Jun 2022 13:01

Instrument :

MSVOA\_N

ClientSampleId :

VN0609WBSD01

Tgt Ion: 56 Resp: 8754

Ion Ratio Lower Upper

56 100

69 34.9 27.8 41.8





84 75.7 76.9 115.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/10/2022

Supervised By :Mahesh Dadoda 06/10/2022



#32

1,1,1-Trichloroethane

Concen: 19.587 ug/l

RT: 8.010 min Scan# 1024

Delta R.T. 0.000 min

Lab File: VN072725.D

Acq: 09 Jun 2022 13:01




Tgt Ion: 97 Resp: 93923


Ion Ratio Lower Upper

97 100

99 64.6 50.6 76.0

61 49.4 36.0 54.0





#33

1,2-Dichloroethane-d4

Concen: 52.572 ug/l

RT: 8.428 min Scan# 1096

Delta R.T. -0.006 min

Lab File: VN072725.D

Acq: 09 Jun 2022 13:01

Instrument :

MSVOA\_N

ClientSampleId :

VN0609WBSD01

Tgt Ion: 65 Resp: 20948

Ion Ratio Lower Upper

65 100





67 49.7 0.0 104.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/10/2022

Supervised By :Mahesh Dadoda 06/10/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.963 min Scan# 1186

Delta R.T. -0.000 min

Lab File: VN072725.D

Acq: 09 Jun 2022 13:01




Tgt Ion:114 Resp: 274397


Ion Ratio Lower Upper

114 100

63 23.5 0.0 40.0


88 16.5 0.0 32.6





#35  
Dibromofluoromethane  
Concen: 51.323 ug/l  
RT: 8.016 min Scan# 1025  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01



Tgt Ion: 113 Resp: 13934  
Ion Ratio Lower Upper  
113 100  
111 101.9 81.4 122.0  
192 14.9 17.0 25.6




Manual Integrations  
APPROVED


Reviewed By :John Carlone 06/10/2022  
Supervised By :Mahesh Dadoda 06/10/2022



#36  
1,1-Dichloropropene  
Concen: 19.275 ug/l  
RT: 8.216 min Scan# 1059  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01


Tgt Ion: 75 Resp: 72643  
Ion Ratio Lower Upper  
75 100  
110 31.6 18.3 54.8  
77 31.0 25.1 37.7





#37  
Ethyl Acetate  
Concen: 20.541 ug/l  
RT: 7.404 min Scan# 91  
Delta R.T. -0.006 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01



Tgt Ion: 43 Resp: 12027  
Ion Ratio Lower Upper  
43 100  
61 13.0 11.8 17.6  
70 9.3 9.3 13.9




Manual Integrations  
APPROVED


Reviewed By :John Carlone 06/10/2022  
Supervised By :Mahesh Dadoda 06/10/2022



#38  
Carbon Tetrachloride  
Concen: 19.918 ug/l  
RT: 8.204 min Scan# 1057  
Delta R.T. 0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01

Tgt Ion:117 Resp: 80220  
Ion Ratio Lower Upper  
117 100  
119 97.3 75.8 113.8  
121 31.1 24.6 37.0





#39

Methylcyclohexane

Concen: 18.689 ug/l

RT: 9.457 min Scan# 1100

Delta R.T. -0.000 min

Lab File: VN072725.D


Acq: 09 Jun 2022 13:01

Instrument :

MSVOA\_N

ClientSampleId :

VN0609WBSD01






|             |                |
|-------------|----------------|
| Tgt Ion: 83 | Resp: 81730    |
| Ion Ratio   | Lower Upper    |
| 83          | 100            |
| 55          | 84.2 57.8 86.8 |
| 98          | 48.2 39.4 59.2 |

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/10/2022

Supervised By :Mahesh Dadoda 06/10/2022



#40

Benzene

Concen: 20.273 ug/l




RT: 8.457 min Scan# 1100

Delta R.T. 0.000 min


Lab File: VN072725.D

Acq: 09 Jun 2022 13:01

|             |                |
|-------------|----------------|
| Tgt Ion: 78 | Resp: 231851   |
| Ion Ratio   | Lower Upper    |
| 78          | 100            |
| 77          | 23.8 19.4 29.0 |








#41  
Methacrylonitrile  
Concen: 29.194 ug/l  
RT: 7.628 min Scan# 911  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01




Tgt Ion: 41 Resp: 57620  
Ion Ratio Lower Upper  
41 100  
39 63.3 48.6 73.0  
67 74.1 73.4 110.2  
52 34.6 27.7 41.5

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/10/2022  
Supervised By :Mahesh Dadoda 06/10/2022




#42  
1,2-Dichloroethane  
Concen: 21.559 ug/l  
RT: 8.522 min Scan# 1111  
Delta R.T. -0.006 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01




Tgt Ion: 62 Resp: 102226  
Ion Ratio Lower Upper  
62 100  
98 8.3 0.0 19.4





#43  
Isopropyl Acetate  
Concen: 21.286 ug/l  
RT: 8.557 min Scan# 1117  
Delta R.T. 0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01




Tgt Ion: 43 Resp: 19410  
Ion Ratio Lower Upper  
43 100  
61 24.1 21.9 32.9  
87 11.9 12.0 18.0

Manual Integrations  
APPROVED



Reviewed By :John Carlone 06/10/2022  
Supervised By :Mahesh Dadoda 06/10/2022




#44  
Trichloroethene  
Concen: 20.142 ug/l  
RT: 9.210 min Scan# 1228  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01



Tgt Ion:130 Resp: 53629  
Ion Ratio Lower Upper  
130 100  
95 104.1 0.0 186.8





#45

1,2-Dichloropropane

Concen: 20.923 ug/l

RT: 9.486 min Scan# 1275

Delta R.T. 0.000 min

Lab File: VN072725.D

Acq: 09 Jun 2022 13:01

Instrument :

MSVOA\_N

ClientSampleId :

VN0609WBSD01

Tgt Ion: 63 Resp: 64660

Ion Ratio Lower Upper

63 100





65 29.5 24.9 37.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/10/2022

Supervised By :Mahesh Dadoda 06/10/2022



#46

Dibromomethane

Concen: 20.656 ug/l

RT: 9.575 min Scan# 1290

Delta R.T. -0.000 min

Lab File: VN072725.D

Acq: 09 Jun 2022 13:01




Tgt Ion: 93 Resp: 44625


Ion Ratio Lower Upper

93 100

95 85.6 66.9 100.3


174 81.7 86.6 130.0#





#47  
Bromodichloromethane  
Concen: 20.958 ug/l  
RT: 9.757 min Scan# 11  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01



Tgt Ion: 83 Resp: 9307  
Ion Ratio Lower Upper  
83 100  
85 60.8 50.5 75.7  
127 8.3 6.8 10.2


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






06/10/2022  
Supervised By :Mahesh  
Dadoda


06/10/2022



#48  
Methyl methacrylate  
Concen: 19.726 ug/l  
RT: 9.557 min Scan# 1287  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01

Tgt Ion: 41 Resp: 85105  
Ion Ratio Lower Upper  
41 100  
69 81.9 74.7 112.1  
39 72.4 45.6 68.4#





#49

1,4-Dioxane

Concen: 439.688 ug/l

RT: 9.563 min Scan# 1289

Delta R.T. -0.006 min

Lab File: VN072725.D

Acq: 09 Jun 2022 13:01

Instrument :

MSVOA\_N



ClientSampleId :

VN0609WBSD01

|           |       |       |      |
|-----------|-------|-------|------|
| Tgt Ion:  | 88    | Resp: | 3467 |
| Ion Ratio | Lower | Upper |      |
| 88        | 100   |       |      |
| 43        | 38.1  | 20.2  | 30.4 |
| 58        | 72.1  | 54.2  | 81.4 |



Manual Integrations  
APPROVED

Reviewed By :John  
Carlone



06/10/2022  
Supervised By :Mahesh  
Dadoda

06/10/2022



#50

Toluene-d8

Concen: 50.591 ug/l




RT: 10.439 min Scan# 1437


Delta R.T. 0.000 min

Lab File: VN072725.D

Acq: 09 Jun 2022 13:01


|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 98    | Resp: | 548043 |
| Ion Ratio | Lower | Upper |        |
| 98        | 100   |       |        |
| 100       | 64.1  | 52.6  | 78.8   |





#51  
4-Methyl-2-Pentanone  
Concen: 106.568 ug/l  
RT: 10.328 min Scan# 1418  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01



Tgt Ion: 43 Resp: 63285  
Ion Ratio Lower Upper  
43 100  
58 36.8 34.1 51.1


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




06/10/2022  
Supervised By :Mahesh  
Dadoda



06/10/2022




#52  
Toluene  
Concen: 20.471 ug/l  
RT: 10.504 min Scan# 1448  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01




Tgt Ion: 92 Resp: 146175  
Ion Ratio Lower Upper  
92 100  
91 175.0 138.6 207.8





#53  
t-1,3-Dichloropropene  
Concen: 21.134 ug/l  
RT: 10.716 min Scan# 1484  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01



Tgt Ion: 75 Resp: 104264  
Ion Ratio Lower Upper  
75 100  
77 32.8 25.6 38.4


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




06/10/2022  
Supervised By :Mahesh  
Dadoda



06/10/2022




#54  
cis-1,3-Dichloropropene  
Concen: 21.488 ug/l  
RT: 10.186 min Scan# 1394  
Delta R.T. 0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01




Tgt Ion: 75 Resp: 109535  
Ion Ratio Lower Upper  
75 100  
77 32.5 25.2 37.8  
39 61.9 35.7 53.5#





#55  
1,1,2-Trichloroethane  
Concen: 20.933 ug/l  
RT: 10.898 min Scan# 1514  
Delta R.T. 0.006 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01



Tgt Ion: 97 Resp: 64504  
Ion Ratio Lower Upper  
97 100  
83 90.3 66.2 99.4  
85 56.0 43.9 65.9  
99 61.8 50.6 76.0


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




06/10/2022  
Supervised By :Mahesh  
Dadoda



06/10/2022




#56  
Ethyl methacrylate  
Concen: 21.649 ug/l  
RT: 10.757 min Scan# 1491  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01



Tgt Ion: 69 Resp: 117985  
Ion Ratio Lower Upper  
69 100  
41 73.2 47.8 71.6#  
39 46.3 27.3 40.9#








#57  
1,3-Dichloropropane  
Concen: 21.226 ug/l  
RT: 11.039 min Scan# 11679  
Delta R.T. 0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01



Tgt Ion: 76 Resp: 11679  
Ion Ratio Lower Upper  
76 100  
78 32.1 26.2 39.2


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






06/10/2022  
Supervised By :Mahesh  
Dadoda


06/10/2022



#58  
2-Chloroethyl Vinyl ether  
Concen: 109.803 ug/l  
RT: 10.039 min Scan# 1369  
Delta R.T. 0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01


Tgt Ion: 63 Resp: 238496  
Ion Ratio Lower Upper  
63 100  
106 26.3 24.6 36.8





#59  
2-Hexanone  
Concen: 109.344 ug/l  
RT: 11.080 min Scan# 1546  
Delta R.T. 0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01



Tgt Ion: 43 Resp: 47748  
Ion Ratio Lower Upper  
43 100  
58 51.6 29.7 89.1


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






06/10/2022  
Supervised By :Mahesh  
Dadoda


06/10/2022



#60  
Dibromochloromethane  
Concen: 21.838 ug/l  
RT: 11.233 min Scan# 1572  
Delta R.T. 0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01


Tgt Ion:129 Resp: 69843  
Ion Ratio Lower Upper  
129 100  
127 75.5 38.8 116.4





#61  
1,2-Dibromoethane  
Concen: 21.675 ug/l  
RT: 11.345 min Scan# 1591  
Delta R.T. 0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01



Tgt Ion:107 Resp: 68450  
Ion Ratio Lower Upper  
107 100  
109 91.9 75.0 112.6


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




06/10/2022  
Supervised By :Mahesh  
Dadoda



06/10/2022




#62  
4-Bromofluorobenzene  
Concen: 51.873 ug/l  
RT: 12.727 min Scan# 1826  
Delta R.T. 0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01




Tgt Ion: 95 Resp: 192823  
Ion Ratio Lower Upper  
95 100  
174 64.2 0.0 167.6  
176 61.1 0.0 163.4





#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.739 min Scan# 117  
Delta R.T. -0.006 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01



Tgt Ion:117 Resp: 29524  
Ion Ratio Lower Upper  
117 100  
82 59.9 44.2 66.2  
119 32.5 25.4 38.0


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




06/10/2022  
Supervised By :Mahesh  
Dadoda



06/10/2022




#64  
Tetrachloroethene  
Concen: 19.063 ug/l  
RT: 10.975 min Scan# 1528  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01




Tgt Ion:164 Resp: 47385  
Ion Ratio Lower Upper  
164 100  
166 134.0 104.2 156.2  
129 102.8 73.9 110.9  
131 101.8 69.4 104.2





#65  
Chlorobenzene  
Concen: 20.334 ug/l  
RT: 11.769 min Scan# 1663  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01



Tgt Ion:112 Resp: 159975  
Ion Ratio Lower Upper  
112 100  
114 31.3 25.4 38.2


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




06/10/2022  
Supervised By :Mahesh  
Dadoda



06/10/2022




#66  
1,1,1,2-Tetrachloroethane  
Concen: 20.175 ug/l  
RT: 11.839 min Scan# 1675  
Delta R.T. 0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01



Tgt Ion:131 Resp: 63140  
Ion Ratio Lower Upper  
131 100  
133 90.2 47.8 143.4  
119 64.3 33.4 100.2







#67  
Ethyl Benzene  
Concen: 19.820 ug/l  
RT: 11.839 min Scan# 1675  
Delta R.T. -0.006 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01

Tgt Ion: 91 Resp: 293200  
Ion Ratio Lower Upper  
91 100  
106 29.4 24.9 37.3



Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






06/10/2022  
Supervised By :Mahesh  
Dadoda


06/10/2022



#68  
m/p-Xylenes  
Concen: 40.224 ug/l  
RT: 11.951 min Scan# 1694  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01

Tgt Ion:106 Resp: 211459  
Ion Ratio Lower Upper  
106 100  
91 214.7 158.9 238.3







#69  
o-Xylene  
Concen: 20.618 ug/l  
RT: 12.280 min Scan# 1749  
Delta R.T. 0.006 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01

Tgt Ion:106 Resp: 108959  
Ion Ratio Lower Upper  
106 100  
91 225.0 106.0 318.0



Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






06/10/2022  
Supervised By :Mahesh  
Dadoda


06/10/2022



#70  
Styrene  
Concen: 20.718 ug/l  
RT: 12.292 min Scan# 1752  
Delta R.T. 0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01


Tgt Ion:104 Resp: 172074  
Ion Ratio Lower Upper  
104 100  
78 57.0 39.5 59.3  
103 55.3 44.0 66.0





#71  
Bromoform  
Concen: 21.218 ug/l  
RT: 12.457 min Scan# 1780  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01



Tgt Ion:173 Resp: 46700  
Ion Ratio Lower Upper  
173 100  
175 46.8 24.6 73.6  
254 0.0 0.1 0.1


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




06/10/2022  
Supervised By :Mahesh  
Dadoda



06/10/2022




#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 13.668 min Scan# 1986  
Delta R.T. -0.001 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01



Tgt Ion:152 Resp: 134056  
Ion Ratio Lower Upper  
152 100  
115 65.6 28.9 86.7  
150 170.7 0.0 356.2








#73  
Isopropylbenzene  
Concen: 19.575 ug/l  
RT: 12.574 min Scan# 1800  
Delta R.T. 0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01



Tgt Ion:105 Resp: 281484  
Ion Ratio Lower Upper  
105 100  
120 25.5 13.5 40.4


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




06/10/2022  
Supervised By :Mahesh  
Dadoda



06/10/2022




#74  
N-ethyl acetate  
Concen: 21.412 ug/l  
RT: 12.386 min Scan# 1768  
Delta R.T. 0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01




Tgt Ion: 43 Resp: 143102  
Ion Ratio Lower Upper  
43 100  
70 38.0 39.0 58.4#  
55 21.8 21.1 31.7  
61 21.8 21.1 31.7





#75  
1,1,2,2-Tetrachloroethane  
Concen: 20.511 ug/l  
RT: 12.821 min Scan# 1842  
Delta R.T. -0.006 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01



Tgt Ion: 83 Resp: 10238  
Ion Ratio Lower Upper  
83 100  
131 10.5 5.3 15.9  
85 64.8 32.3 96.9


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






06/10/2022  
Supervised By :Mahesh  
Dadoda


06/10/2022



#76  
1,2,3-Trichloropropane  
Concen: 23.730 ug/l m  
RT: 12.874 min Scan# 1851  
Delta R.T. 0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01


Tgt Ion: 75 Resp: 102654  
Ion Ratio Lower Upper  
75 100  
77 152.8 98.5 295.5





#77  
Bromobenzene  
Concen: 20.290 ug/l  
RT: 12.857 min Scan# 1848  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01



Tgt Ion:156 Resp: 5751  
Ion Ratio Lower Upper  
156 100  
77 272.7 100.5 301.5  
158 97.2 48.9 146.6


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






06/10/2022  
Supervised By :Mahesh  
Dadoda


06/10/2022



#78  
n-propylbenzene  
Concen: 20.246 ug/l  
RT: 12.916 min Scan# 1858  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01


Tgt Ion: 91 Resp: 317860  
Ion Ratio Lower Upper  
91 100  
120 20.8 11.9 35.5





#79  
2-Chlorotoluene  
Concen: 19.899 ug/l  
RT: 13.004 min Scan# 1873  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01



Tgt Ion: 91 Resp: 19598  
Ion Ratio Lower Upper  
91 100  
126 31.1 17.3 51.9


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






06/10/2022  
Supervised By :Mahesh  
Dadoda


06/10/2022



#80  
1,3,5-Trimethylbenzene  
Concen: 20.443 ug/l  
RT: 13.057 min Scan# 1882  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01


Tgt Ion:105 Resp: 232925  
Ion Ratio Lower Upper  
105 100  
120 47.2 24.9 74.7





#81  
trans-1,4-Dichloro-2-butene  
Concen: 19.723 ug/l  
RT: 12.621 min Scan# 1808  
Delta R.T. -0.001 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01



Tgt Ion: 75 Resp: 32624  
Ion Ratio Lower Upper  
75 100  
53 119.1 78.0 117.0  
89 47.3 37.2 55.8


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




06/10/2022  
Supervised By :Mahesh  
Dadoda



06/10/2022




#82  
4-Chlorotoluene  
Concen: 20.434 ug/l  
RT: 13.104 min Scan# 1890  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01




Tgt Ion: 91 Resp: 185879  
Ion Ratio Lower Upper  
91 100  
126 30.8 17.2 51.6





#83  
tert-Butylbenzene  
Concen: 19.813 ug/l  
RT: 13.321 min Scan# 1927  
Delta R.T. 0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01



Tgt Ion:119 Resp: 198274  
Ion Ratio Lower Upper  
119 100  
91 75.6 31.6 94.8  
134 26.5 13.7 41.1


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




06/10/2022  
Supervised By :Mahesh  
Dadoda



06/10/2022




#84  
1,2,4-Trimethylbenzene  
Concen: 20.727 ug/l  
RT: 13.363 min Scan# 1934  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01




Tgt Ion:105 Resp: 225961  
Ion Ratio Lower Upper  
105 100  
120 42.7 23.3 69.8





#85  
 sec-Butylbenzene  
 Concen: 20.395 ug/l  
 RT: 13.498 min Scan# 1957  
 Delta R.T. -0.000 min  
 Lab File: VN072725.D  
 Acq: 09 Jun 2022 13:01


Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0609WBSD01



Tgt Ion:105 Resp: 275395  
 Ion Ratio Lower Upper  
 105 100  
 134 19.0 10.4 31.1


Manual Integrations  
 APPROVED

Reviewed By :John  
 Carlone






06/10/2022  
 Supervised By :Mahesh  
 Dadoda


06/10/2022



#86  
 p-Isopropyltoluene  
 Concen: 20.331 ug/l  
 RT: 13.610 min Scan# 1976  
 Delta R.T. -0.000 min  
 Lab File: VN072725.D  
 Acq: 09 Jun 2022 13:01


Tgt Ion:119 Resp: 215953  
 Ion Ratio Lower Upper  
 119 100  
 134 26.1 13.3 39.8  
 91 27.1 11.3 33.9





#87  
1,3-Dichlorobenzene  
Concen: 20.558 ug/l  
RT: 13.616 min Scan# 1976  
Delta R.T. 0.006 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01



Tgt Ion:146 Resp: 98980  
Ion Ratio Lower Upper  
146 100  
111 44.6 20.0 60.0  
148 64.1 32.0 96.0


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




06/10/2022  
Supervised By :Mahesh  
Dadoda



06/10/2022




#88  
1,4-Dichlorobenzene  
Concen: 20.723 ug/l  
RT: 13.692 min Scan# 1990  
Delta R.T. 0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01



Tgt Ion:146 Resp: 97747  
Ion Ratio Lower Upper  
146 100  
111 44.9 19.1 57.1  
148 63.2 31.4 94.3








#89  
n-Butylbenzene  
Concen: 20.340 ug/l  
RT: 13.939 min Scan# 2032  
Delta R.T. 0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01



Tgt Ion: 91 Resp: 182025  
Ion Ratio Lower Upper  
91 100  
92 52.3 26.9 80.7  
134 23.8 13.6 40.8


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




06/10/2022  
Supervised By :Mahesh  
Dadoda



06/10/2022




#90  
Hexachloroethane  
Concen: 18.950 ug/l  
RT: 14.204 min Scan# 2077  
Delta R.T. -0.006 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01




Tgt Ion:117 Resp: 43566  
Ion Ratio Lower Upper  
117 100  
201 59.3 40.6 122.0





#91  
1,2-Dichlorobenzene  
Concen: 20.791 ug/l  
RT: 13.986 min Scan# 2040  
Delta R.T. 0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01



Tgt Ion:146 Resp: 98114  
Ion Ratio Lower Upper  
146 100  
111 48.0 20.9 62.8  
148 65.5 32.1 96.3


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






06/10/2022  
Supervised By :Mahesh  
Dadoda


06/10/2022



#92  
1,2-Dibromo-3-Chloropropane  
Concen: 19.589 ug/l  
RT: 14.598 min Scan# 2144  
Delta R.T. -0.006 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01


Tgt Ion: 75 Resp: 23619  
Ion Ratio Lower Upper  
75 100  
155 64.4 44.9 134.5  
157 82.6 59.0 176.8





#93  
1,2,4-Trichlorobenzene  
Concen: 21.034 ug/l  
RT: 15.262 min Scan# 21  
Delta R.T. -0.001 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01



Tgt Ion:180 Resp: 4775  
Ion Ratio Lower Upper  
180 100  
182 89.9 47.5 142.5  
145 35.2 14.5 43.5


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




06/10/2022  
Supervised By :Mahesh  
Dadoda



06/10/2022




#94  
Hexachlorobutadiene  
Concen: 21.204 ug/l  
RT: 15.368 min Scan# 2275  
Delta R.T. 0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01




Tgt Ion:225 Resp: 21563  
Ion Ratio Lower Upper  
225 100  
223 61.7 32.4 97.2  
227 66.1 31.8 95.4





#95  
Naphthalene  
Concen: 20.795 ug/l  
RT: 15.504 min Scan# 2298  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0609WBSD01



Tgt Ion:128 Resp: 18316  
Ion Ratio Lower Upper  
128 100  
127 13.2 10.3 15.5  
129 11.0 8.6 13.0


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




06/10/2022  
Supervised By :Mahesh  
Dadoda

06/10/2022



#96  
1,2,3-Trichlorobenzene  
Concen: 21.991 ug/l  
RT: 15.698 min Scan# 2331  
Delta R.T. -0.000 min  
Lab File: VN072725.D  
Acq: 09 Jun 2022 13:01



Tgt Ion:180 Resp: 49714  
Ion Ratio Lower Upper  
180 100  
182 92.6 47.8 143.3  
145 35.2 15.5 46.5

