

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060923\
 Data File : VN078140.D
 Acq On : 09 Jun 2023 12:46
 Operator : JC\MD
 Sample : 03123-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW11-GW-678-680

Quant Time: Jun 09 23:06:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/12/2023
 Supervised By : Mahesh Dadoda 06/12/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	452302	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	823324	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	719785	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	218294	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	312012	51.738	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	= 103.480%		
35) Dibromofluoromethane	8.171	113	270099	51.885	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	= 103.760%		
50) Toluene-d8	10.571	98	982558	48.960	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	= 97.920%		
62) 4-Bromofluorobenzene	12.853	95	290777	46.607	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery	= 93.220%		
Target Compounds						Qvalue
16) Acetone	4.459	43	4944	2.848	ug/l #	82
25) 2-Butanone	7.506	43	4205m	1.575	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060923\
Data File : VN078140.D
Acq On : 09 Jun 2023 12:46
Operator : JC\MD
Sample : 03123-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW11-GW-678-680

Quant Time: Jun 09 23:06:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 06/12/2023
Supervised By :Mahesh Dadoda 06/12/2023

