

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060923\
 Data File : VN078143.D
 Acq On : 09 Jun 2023 14:11
 Operator : JC\MD
 Sample : 03123-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW11-GW-718-720

Quant Time: Jun 09 23:07:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	416697	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	768570	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	675607	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	204448	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	294463	53.001	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.000%
35) Dibromofluoromethane	8.177	113	256801	52.844	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.680%
50) Toluene-d8	10.571	98	923214	49.280	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.560%
62) 4-Bromofluorobenzene	12.853	95	281861	48.397	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060923\
Data File : VN078143.D
Acq On : 09 Jun 2023 14:11
Operator : JC\MD
Sample : 03123-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW11-GW-718-720

Quant Time: Jun 09 23:07:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

