

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060923\
 Data File : VN078151.D
 Acq On : 09 Jun 2023 17:22
 Operator : JC\MD
 Sample : 03119-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Jun 09 23:12:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	438224	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	799372	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	699420	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	217702	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	307044	52.550	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.100%
35) Dibromofluoromethane	8.171	113	268783	53.179	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.360%
50) Toluene-d8	10.571	98	962809	49.413	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.820%
62) 4-Bromofluorobenzene	12.853	95	289043	47.717	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060923\
Data File : VN078151.D
Acq On : 09 Jun 2023 17:22
Operator : JC\MD
Sample : 03119-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Jun 09 23:12:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

