

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060923\
 Data File : VN078153.D
 Acq On : 09 Jun 2023 18:10
 Operator : JC\MD
 Sample : 03119-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Jun 09 23:13:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	422209	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	773602	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	667068	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	197276	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	300429	53.368	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.740%	
35) Dibromofluoromethane	8.177	113	261485	53.458	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.920%	
50) Toluene-d8	10.571	98	942980	50.007	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.020%	
62) 4-Bromofluorobenzene	12.853	95	270866	46.206	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.420%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060923\
Data File : VN078153.D
Acq On : 09 Jun 2023 18:10
Operator : JC\MD
Sample : 03119-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Jun 09 23:13:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

