

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060925\
 Data File : VN086894.D
 Acq On : 09 Jun 2025 11:12
 Operator : JC\MD
 Sample : Q2254-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-182-GW-810-812

Quant Time: Jun 09 13:27:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	313915	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	549826	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	462153	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	222690	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	179677	42.750	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	85.500%		
35) Dibromofluoromethane	8.177	113	157362	48.294	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.580%		
50) Toluene-d8	10.571	98	644850	49.992	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	99.980%		
62) 4-Bromofluorobenzene	12.847	95	229953	47.983	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	95.960%		
Target Compounds					Qvalue	
16) Acetone	4.459	43	4943	2.218	ug/l	96
52) Toluene	10.629	92	7042	0.726	ug/l	93
68) m/p-Xylenes	12.070	106	2159	0.321	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060925\
Data File : VN086894.D
Acq On : 09 Jun 2025 11:12
Operator : JC\MD
Sample : Q2254-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-182-GW-810-812

Quant Time: Jun 09 13:27:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

