

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060925\
 Data File : VN086895.D
 Acq On : 09 Jun 2025 11:33
 Operator : JC\MD
 Sample : Q2216-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 3865

Quant Time: Jun 09 13:28:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	391760	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	696305	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	603798	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	294406	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	226237	43.132	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.260%
35) Dibromofluoromethane	8.177	113	197074	47.759	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.520%
50) Toluene-d8	10.565	98	837075	51.242	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.480%
62) 4-Bromofluorobenzene	12.847	95	298237	49.140	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	98.280%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.542	59	22113	15.537	ug/l	99
16) Acetone	4.453	43	6650	2.391	ug/l	94
20) Methylene Chloride	5.295	84	1454	0.279	ug/l	87
52) Toluene	10.630	92	2555	0.208	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060925\
Data File : VN086895.D
Acq On : 09 Jun 2025 11:33
Operator : JC\MD
Sample : Q2216-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
3865

Quant Time: Jun 09 13:28:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

