

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061019\
 Data File : VN056157.D
 Acq On : 10 Jun 2019 9:56
 Operator : JC/SP
 Sample : VN0610WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0610WBL01

Quant Time: Jun 11 03:19:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	292845	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	443159	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	412863	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	121780	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	152404	52.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.82%	
35) Dibromofluoromethane	7.59	113	143434	52.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.50%	
50) Toluene-d8	10.09	98	548422	52.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.84%	
62) 4-Bromofluorobenzene	12.40	95	162457	45.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.32%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061019\
Data File : VN056157.D
Acq On : 10 Jun 2019 9:56
Operator : JC/SP
Sample : VN0610WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0610WBL01

Quant Time: Jun 11 03:19:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

