

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061019\
 Data File : VN056167.D
 Acq On : 10 Jun 2019 14:07
 Operator : JC/SP
 Sample : K3255-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE117D2-20190603

Quant Time: Jun 11 03:58:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	295899	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	457822	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	446012	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	143076	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	162379	55.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.54%	
35) Dibromofluoromethane	7.59	113	146824	52.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.54%	
50) Toluene-d8	10.09	98	589036	54.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.00%	
62) 4-Bromofluorobenzene	12.40	95	190784	51.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.80%	
Target Compounds						
44) Trichloroethene	8.83	130	1515	0.441	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061019\
Data File : VN056167.D
Acq On : 10 Jun 2019 14:07
Operator : JC/SP
Sample : K3255-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE117D2-20190603

Quant Time: Jun 11 03:58:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

