

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061019\
 Data File : VN056171.D
 Acq On : 10 Jun 2019 15:47
 Operator : JC/SP
 Sample : K3280-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-FW-MW02-20190608

Quant Time: Jun 11 12:04:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	314621	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	472685	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	460998	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	140478	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	164766	52.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.48%	
35) Dibromofluoromethane	7.59	113	151049	52.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.16%	
50) Toluene-d8	10.09	98	600409	53.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.60%	
62) 4-Bromofluorobenzene	12.40	95	186425	49.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.24%	
Target Compounds						
44) Trichloroethene	8.83	130	6045	1.704	ug/l	96
52) Toluene	10.15	92	2886	0.376	ug/l	96
64) Tetrachloroethene	10.63	164	12114	3.036	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061019\
Data File : VN056171.D
Acq On : 10 Jun 2019 15:47
Operator : JC/SP
Sample : K3280-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-FW-MW02-20190608

Quant Time: Jun 11 12:04:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

