

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061019\
 Data File : VN056173.D
 Acq On : 10 Jun 2019 16:38
 Operator : JC/SP
 Sample : K3280-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-HN-MW271-20190608

Quant Time: Jun 11 04:12:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	316905	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	476338	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	448383	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	128733	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	164565	52.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.60%	
35) Dibromofluoromethane	7.59	113	151796	51.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.88%	
50) Toluene-d8	10.09	98	599280	53.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.58%	
62) 4-Bromofluorobenzene	12.40	95	179059	46.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.64%	
Target Compounds						
16) Acetone	3.82	43	3024	2.009	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061019\
Data File : VN056173.D
Acq On : 10 Jun 2019 16:38
Operator : JC/SP
Sample : K3280-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-HN-MW271-20190608

Quant Time: Jun 11 04:12:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

