

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061019\
 Data File : VN056174.D
 Acq On : 10 Jun 2019 17:03
 Operator : JC/SP
 Sample : K3280-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-HN-MW29IR-20190608

Quant Time: Jun 11 04:14:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	313506	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	481162	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	458889	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	139235	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	169989	54.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.22%	
35) Dibromofluoromethane	7.59	113	157502	53.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.70%	
50) Toluene-d8	10.09	98	608675	53.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.16%	
62) 4-Bromofluorobenzene	12.40	95	187317	48.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.98%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061019\
Data File : VN056174.D
Acq On : 10 Jun 2019 17:03
Operator : JC/SP
Sample : K3280-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-HN-MW29IR-20190608

Quant Time: Jun 11 04:14:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

